




THE FORMULATION ENGINEER

cGMP Processing Equipment for
**Pharmaceutical & Nutraceutical
Manufacturing**

Mixers • Mills • Extruders • Bulk Material Handling

*From API Micronization to Solid Dosage Coating — The Complete
Technical Reference*

www.procxinc.com | john.paul@procxinc.com



Preface: Precision at Every Step

In pharmaceutical and nutraceutical manufacturing, the processing equipment is not a commodity — it is a regulated system. Every mixer, every mill, every extruder that comes into contact with a drug product or dietary supplement is a critical quality system element, subject to qualification, validation, and ongoing monitoring. The performance envelope of the equipment defines the performance envelope of the product. Get the processing right, and every tablet, every capsule, every sterile vial meets its specification. Get it wrong, and the consequences range from product recall to patient harm.

At Proc-X Manufacturing Group, we have spent decades working at this intersection — the place where materials science meets mechanical engineering, where FDA regulations meet real-world manufacturing. We have designed mixing systems for high-potency APIs with occupational exposure limits (OELs) in the nanogram-per-cubic-meter range. We have built jet milling systems for APIs that cannot tolerate a single degree of temperature rise. We have integrated continuous manufacturing lines that compress weeks of batch processing into hours of continuous, fully monitored production.

This book is our complete technical reference for pharmaceutical and nutraceutical processing. It covers the science of granulation, the engineering of fluid bed systems, the pharmacopeial requirements for API particle size, the cGMP implications of equipment surface finish and cleanability, and the regulatory framework — 21 CFR Part 211, 21 CFR Part 11, ICH Q8/Q9/Q10, and the emerging ICH Q13 continuous manufacturing guideline — that governs every decision in this industry.

We have organized it around how the industry actually works: from raw API to finished dosage form, from small-scale R&D to full commercial production, from traditional batch manufacturing to the continuous processing frontier. The Proc-X equipment that appears throughout these pages is designed to serve every phase of this journey — and to make compliance not a constraint but a built-in feature of the process.

The Engineering Team
Proc-X Manufacturing Group

Table of Contents

Open in Microsoft Word, press Ctrl+A, then F9 to generate the live Table of Contents with accurate page numbers.

Preface: Precision at Every Step	2
Table of Contents	3
Chapter 1: Proc-X and the Pharmaceutical Processing World	9
1.1 Why Pharmaceutical Processing Is Different	9
1.2 Core Equipment Competencies in Pharmaceutical Service	9
1.3 cGMP Design Philosophy at Proc-X.....	10
Chapter 2: The Pharmaceutical & Nutraceutical Manufacturing Landscape	11
2.1 Scale and Structure of the Industry.....	11
2.2 The Four Application Domains.....	11
Solid Dosage Forms.....	11
API Processing.....	11
Nutraceuticals & Supplements.....	11
Sterile & Injectable	12
2.3 Dosage Form Processing Chains	12
2.4 Material Classification Systems Relevant to Processing.....	12
Chapter 3: 21 CFR Part 211, Part 111 & the cGMP Framework for Processing Equipment	14
3.1 21 CFR Part 211 — Pharmaceutical cGMP Equipment Requirements.....	14
211.63 — Equipment Design, Size, and Location.....	14
211.65 — Equipment Construction	14
211.67 — Equipment Cleaning and Maintenance.....	15
211.68 — Automatic, Mechanical, and Electronic Equipment.....	15
3.2 21 CFR Part 111 — Dietary Supplement cGMP	15
3.3 EU GMP Annex 11 and ICH Q10.....	16
3.4 How Equipment Design Prevents Regulatory Findings	16
Chapter 4: Granulation Science — Wet, Dry, and Continuous	18
4.1 Why Granulation?.....	18
4.2 Wet Granulation Mechanisms.....	18
4.3 Dry Granulation and Roller Compaction	19

4.4 Continuous Granulation	19
Chapter 5: High-Shear Wet Granulation Equipment.....	21
5.1 Equipment Architecture	21
5.2 Power and Torque as Process Endpoints	21
5.3 Scale-Up Principle: Impeller Tip Speed.....	22
Chapter 6: Fluid Bed Processing — Granulation, Drying & Coating	23
6.1 The Fluid Bed as a Multi-Function Platform	23
6.2 Top-Spray Granulation Process Parameters.....	23
6.3 Wurster (Bottom-Spray) Coating	24
Chapter 7: Powder Blending for Solid Dosage Manufacturing	25
7.1 Blending Mechanisms and Equipment Selection	25
7.2 Bin Blenders for OEB 3–5 Processing	25
7.3 Blend Endpoint Determination	26
Chapter 8: Film Coating and Pan Coating Systems	27
8.1 Why Coat Tablets?	27
8.2 Side-Vented Coating Pan.....	27
Chapter 9: Twin-Screw Extrusion and Hot-Melt Extrusion in Pharma.....	29
9.1 Hot-Melt Extrusion and Bioavailability Enhancement.....	29
9.2 TSE Screw Design for HME	29
Chapter 10: The API Particle Size Challenge	32
10.1 Why Particle Size Matters	32
10.2 The Noyes-Whitney Equation in Practice	32
10.3 Polymorphism and Milling	33
Chapter 11: Jet Milling and Micronization of APIs	34
11.1 Jet Milling Mechanism.....	34
11.2 Nitrogen Gas Systems for API Milling.....	34
11.3 Scale-Up for API Jet Milling.....	35
Chapter 12: Wet Milling, Nanosizing, and Media Milling	36
12.1 Wet Milling vs. Jet Milling	36
12.2 Nanosizing Technology	36
Chapter 13: Cryogenic Milling for Thermolabile APIs	38
13.1 The Thermolabile Milling Challenge	38
Chapter 14: Continuous API Manufacturing and PAT Integration.....	39

14.1 ICH Q13 and Continuous Manufacturing	39
14.2 PAT in Continuous API Processing.....	39
Chapter 15: The Nutraceutical Processing Landscape.....	42
15.1 The Nutraceutical Market and Its Processing Demands.....	42
15.2 21 CFR Part 111 Processing Requirements	42
Chapter 16: Spray Drying for Nutraceutical Powder Production	44
16.1 The Spray Drying Process	44
16.2 Microencapsulation for Sensitive Ingredients	44
Chapter 17: Botanical Extract Processing and Standardization	46
17.1 The Natural Variability Challenge.....	46
17.2 Milling Botanical Extracts.....	46
Chapter 18: Encapsulation, Softgel & Delivery System Processing.....	48
18.1 Hard Capsule Filling.....	48
18.2 Softgel Encapsulation.....	48
Chapter 19: Aseptic Processing Principles and Equipment.....	51
19.1 Why Sterile Manufacturing Requires a Different Approach.....	51
19.2 Sterile Mixing Equipment.....	51
19.3 Single-Use vs. Stainless Steel for Sterile Mixing	52
Chapter 20: Sterile API Milling and Micronization	53
20.1 The Sterile Milling Challenge.....	53
20.2 DPI API Micronization	53
Chapter 21: Lyophilization Preparation — Formulation Mixing and Filling	54
21.1 Why Lyophilize?	54
21.2 Fill Accuracy Requirements.....	54
Chapter 22: High-Shear Granulators and Fluid Bed Processors	56
22.1 High-Shear Granulator Design Anatomy.....	56
22.2 Fluid Bed Processor Control Architecture.....	56
Chapter 23: Pharmaceutical Mills — From Comil to Jet Mill	58
23.1 The Pharmaceutical Mill Taxonomy.....	58
23.2 Conical Mills — the Pharma Workhorse.....	58
23.3 Mill Selection Decision Logic	58
Chapter 24: Twin-Screw Extruders and Continuous Processing Equipment	60
24.1 TSE Design for Pharmaceutical Applications	60

24.2 Continuous Wet Granulation by TSE	60
Chapter 25: Bulk Material Handling in Pharmaceutical Facilities.....	62
25.1 Contained Powder Transfer for High-Potency APIs.....	62
25.2 Vacuum Conveying in Pharmaceutical Facilities.....	62
25.3 Loss-in-Weight Feeders for Continuous Manufacturing.....	62
Chapter 26: 21 CFR Part 211 — Complete cGMP Compliance for Processing Equipment	65
26.1 Complete 21 CFR Part 211 Equipment Requirements.....	65
Section 211.63 — Equipment Design, Size, and Location	65
Section 211.65 — Equipment Construction.....	65
Section 211.67 — Equipment Cleaning and Maintenance	66
Section 211.68 — Automatic, Mechanical, and Electronic Equipment	66
Section 211.182 — Equipment Cleaning and Use Log	66
26.2 The QbD Framework and Equipment Specification	66
26.3 The Compliance Documentation Package.....	67
Chapter 27: Process Validation — Stage 1, Stage 2, and Stage 3	68
27.1 The Three-Stage Validation Lifecycle.....	68
Stage 1 — Process Design	68
Stage 2 — Process Performance Qualification (PPQ)	68
Stage 3 — Continued Process Verification (CPV).....	68
Chapter 28: Data Integrity, 21 CFR Part 11, and EU Annex 11	70
28.1 The Data Integrity Imperative.....	70
28.2 ALCOA+ in Processing Equipment Systems.....	70
28.3 21 CFR Part 11 Requirements for Electronic Records.....	71
28.4 Common Data Integrity Failures in Processing Equipment.....	71
Chapter 29: Continuous Manufacturing and ICH Q13.....	72
29.1 Continuous Manufacturing — the Paradigm Shift	72
29.2 The Complete Continuous Solid Dose Line.....	72
29.3 Real-Time Release Testing (RTRT).....	73
Chapter 30: The Digital Pharmaceutical Facility — Data, AI, and the Future	74
30.1 Industry 4.0 and Pharma's Unique Regulatory Context	74
30.2 The Pharmaceutical Digital Twin.....	74
30.3 Machine Learning for Pharmaceutical Processing.....	75

APPENDIX A: Glossary of Key Terms 76

APPENDIX B: Equipment Selection Guide..... 78

APPENDIX C: Process Parameter Reference Tables..... 79

 C.1 High-Shear Wet Granulation Reference Parameters 79

 C.2 Jet Milling Reference Parameters for Common API Classes..... 79

 C.3 Fluid Bed Drying Reference Parameters 79

APPENDIX D: Regulatory Standards Index..... 80

APPENDIX E: Proc-X Pharmaceutical Equipment Quick-Reference Index..... 82

PART I

FOUNDATIONS

The Science, the Industry, and the Regulatory Framework

Chapter 1: Proc-X and the Pharmaceutical Processing World

IN THIS CHAPTER

- ▶ Why pharmaceutical processing requires a different engineering philosophy
- ▶ Proc-X's four equipment competencies and how each maps to pharma applications
- ▶ The cGMP design philosophy built into every Proc-X machine
- ▶ Key pharma processing concepts: CQA, CPP, design space, and the QbD framework

1.1 Why Pharmaceutical Processing Is Different

Every industry that uses mixers, mills, and extruders has its own performance requirements. Coatings must be smooth. Foods must be palatable. Plastics must be dimensionally stable. But pharmaceutical manufacturing has a requirement that supersedes all others: the product must be safe and effective for human consumption, every time, without exception. A tablet with 95% of label-claimed active ingredient is not a slightly inferior tablet — it is a potentially harmful drug product that could fail to treat a patient's condition. A tablet with 105% is equally problematic in the opposite direction.

This fundamental requirement — that every unit of every batch meets its specification — shapes everything in pharmaceutical manufacturing, starting with the equipment. It explains why pharmaceutical mixers are equipped with sensors that validate blend uniformity rather than simply mixing for a fixed time. It explains why pharmaceutical mills are specified not merely for throughput but for the particle size distribution they produce, measured to a pharmacopeial standard. It explains why every surface that contacts the product must be smooth enough to clean, inert enough not to react with the drug, and documented well enough to prove it.

Proc-X builds equipment for this environment. We bring the same engineering rigor to pharmaceutical processing that the industry applies to its products.

1.2 Core Equipment Competencies in Pharmaceutical Service

- **Mixers:** High-shear granulators, fluid bed processors, V-blenders, bin blenders, ribbon blenders, and tumble blenders. From first-in-human clinical batches (1–5 kg) to commercial production (500+ kg). cGMP construction standard — 316L SS product contact surfaces, $R_a \leq 0.8 \mu\text{m}$ finish, validated cleaning protocols.
- **Mills:** Jet mills, pin mills, hammer mills, conical mills (Comils), and oscillating granulators. Particle size targets from $D_{90} < 200 \mu\text{m}$ (milling for blend uniformity) to $D_{90} < 5 \mu\text{m}$ (micronization of BCS Class II APIs for bioavailability enhancement). Cryogenic milling capability for thermolabile APIs.

- Extruders: Twin-screw extruders for hot-melt extrusion (HME), co-extrusion for layered dosage forms, and continuous wet granulation extruders. From laboratory 11 mm systems through production 50–75 mm lines.
- Bulk Material Handling: Contained powder transfer (high-potency APIs), loss-in-weight feeders for continuous manufacturing, vacuum conveying systems, split butterfly valves, and contained charging/discharging for OEB 4/5 materials.

REGULATORY CONTEXT *Quality by Design (QbD) is the modern framework for pharmaceutical product and process development, formalized in ICH Q8 (Pharmaceutical Development), Q9 (Quality Risk Management), and Q10 (Pharmaceutical Quality System). Under QbD, process parameters are linked to product quality through a structured design space. Proc-X equipment — with its comprehensive instrumentation, process monitoring, and electronic data recording — is purpose-built to generate the data needed to define and demonstrate this design space.*

1.3 cGMP Design Philosophy at Proc-X

cGMP (current Good Manufacturing Practice) requirements under 21 CFR Part 211 impose specific obligations on equipment used in pharmaceutical manufacturing. Proc-X translates these obligations into equipment design standards:

- Surface finish: All product-contact surfaces 316L stainless steel, electropolished to $Ra \leq 0.5 \mu\text{m}$ (or $Ra \leq 0.8 \mu\text{m}$ for less critical surfaces). Documented with profilometer measurement certificates for each surface.
- Cleanability: Equipment designed for CIP (Clean-in-Place) or WIP (Wash-in-Place) with no dead legs, no crevices, and no product-retention areas. Validated cleaning procedures included in the IQ documentation package.
- Containment: Dedicated contained charging and discharging systems for OEB 4 (EL 1–10 $\mu\text{g}/\text{m}^3$) and OEB 5 (EL < 1 $\mu\text{g}/\text{m}^3$) materials. Single-use contact parts available for clinical-phase manufacturing.
- Documentation: Full material traceability, weld certificates, surface finish certificates, calibration records, and design documentation provided with every system.

PROC-X EQUIPMENT SPOTLIGHT Proc-X pharmaceutical equipment is designed from the ground up to meet 21 CFR Part 211.65 and 211.67 — the equipment construction and cleaning requirements. This is not a post-design consideration. Every Proc-X pharmaceutical mixer, mill, and extruder begins with a cGMP Design Specification reviewed against the relevant regulatory requirements before fabrication starts.

Chapter 2: The Pharmaceutical & Nutraceutical Manufacturing Landscape

IN THIS CHAPTER

- The scale and structure of global pharmaceutical and nutraceutical manufacturing
- The four major processing application domains covered in this book
- Dosage form taxonomy: solid oral, sterile, semi-solid, and liquid
- The nutraceutical regulatory landscape vs. pharmaceutical — key distinctions
- Key material classifications: BCS, OEB, DEA schedule

2.1 Scale and Structure of the Industry

The global pharmaceutical industry exceeded \$1.4 trillion in annual revenues, with the nutraceutical and dietary supplement market adding another \$400+ billion. The manufacturing infrastructure behind these products is vast — thousands of manufacturing sites worldwide, ranging from WHO-certified generics facilities in India and China to FDA-inspected innovator plants in the United States, Europe, and Japan. What unites all of them is the need for reliable, validated processing equipment.

The nutraceutical industry, while subject to FDA oversight under 21 CFR Part 111 (dietary supplements cGMP) rather than the pharmaceutical Part 211, shares many of the same processing challenges — and increasingly adopts pharmaceutical-grade equipment and practices as retailers and consumers demand higher quality standards.

2.2 The Four Application Domains

Solid Dosage Forms

Tablets and capsules represent approximately 65% of all pharmaceutical dosage units manufactured globally. The processing sequence — granulation, blending, compression or filling, and coating — involves more processing equipment than any other dosage form category. Proc-X serves every step of this sequence.

API Processing

Active pharmaceutical ingredients (APIs) are the pharmacologically active components of drug products. API particle size, polymorphic form, and surface area directly affect bioavailability, dissolution rate, and content uniformity. API milling and micronization are among the most technically demanding and highly regulated processing operations in the pharmaceutical industry.

Nutraceuticals & Supplements

Botanical extracts, vitamins, minerals, omega fatty acids, probiotics, and sports nutrition products require processing equipment that can handle extraordinary diversity — from

hygroscopic vitamin powders to oil-based omega-3 concentrates to heat-sensitive probiotic cultures. Spray drying, encapsulation, and powder blending are the core processing operations.

Sterile & Injectable

Sterile drug products — injectables, ophthalmics, inhalation solutions — require the highest levels of processing control and environmental segregation. Aseptic mixing, sterile API micronization, lyophilization, and fill-finish are the key processing areas where Proc-X equipment plays a role.

2.3 Dosage Form Processing Chains

Dosage Form	Key Processing Steps	Critical Quality Attributes	Primary Proc-X Equipment
Immediate release tablet	Granulation → Blend → Compress → Coat	Dissolution, content uniformity, hardness	High-shear granulator, fluid bed, bin blender
Modified release tablet	Granulation → Blend → Compress → Film coat (functional)	Release profile, dissolution, coat uniformity	Fluid bed coater, Wurster coater
Hard capsule (HPMC/gelatin)	Blend → Fill → Seal	Content uniformity, disintegration, flow	Bin blender, V-blender, capsule filler
Softgel capsule	API dispersion in oil → Encapsulate	Content uniformity, moisture, leakage	High-shear disperser, encapsulation line
Dry powder inhalation (DPI)	Micronize API → Blend with lactose carrier	Aerodynamic particle size (MMAD 1–5 µm)	Jet mill, fluidized bed blender
Parenteral solution	Dissolve → Filter → Fill → Sterilize	Particulate matter, sterility, pH	Pharmaceutical-grade mixer, sterile filter
Lyophilized injectable	Formulate → Fill → Freeze dry	Residual moisture, cake appearance, reconstitution	Precision filling/mixing, lyophilizer
Dietary supplement capsule	Extract → Standardize → Blend → Fill	Marker compound content, moisture	Spray dryer, V-blender, bin blender

2.4 Material Classification Systems Relevant to Processing

Three classification systems directly influence equipment specification in pharmaceutical processing:

- **BCS Classification:** The Biopharmaceutics Classification System classifies APIs by aqueous solubility and intestinal permeability. BCS Class II (high permeability, low solubility) APIs — which represent ~40% of new chemical entities — require particle size reduction to achieve adequate bioavailability. Jet milling or wet milling to $D_{90} < 10 \mu\text{m}$ is routinely required for BCS Class II APIs. Proc-X jet milling systems are the standard solution.
- **Occupational Exposure Banding (OEB):** APIs are classified by their occupational exposure limit (OEL) from OEB 1 ($\text{OEL} > 1 \text{ mg/m}^3$, standard industrial hygiene) to

OEB 5 (OEL < 0.1 µg/m³, maximum containment). OEB 4 and OEB 5 materials require fully contained processing equipment with < 1 µg/m³ operator exposure, verified by air sampling during contained transfer operations.

- DEA Scheduling: Controlled substances (opioids, stimulants, benzodiazepines) require DEA-registered manufacturing facilities with physical security and accountability systems. Proc-X contained handling systems support controlled substance processing under DEA requirements.

Chapter 3: 21 CFR Part 211, Part 111 & the cGMP Framework for Processing Equipment

IN THIS CHAPTER

- ▶ What 21 CFR Part 211 actually requires of processing equipment — verbatim analysis
- ▶ 21 CFR Part 111 — the dietary supplement cGMP standard and how it mirrors Part 211
- ▶ Equipment qualification under 21 CFR 211.68 — computerized systems
- ▶ EU GMP Annex 11 and ICH Q10 — international harmonization
- ▶ How equipment design decisions prevent 483 observations and warning letters

3.1 21 CFR Part 211 — Pharmaceutical cGMP Equipment Requirements

Title 21 of the Code of Federal Regulations, Part 211 (21 CFR 211) — Current Good Manufacturing Practice for Finished Pharmaceuticals — is the primary U.S. regulatory standard governing pharmaceutical manufacturing. Subpart D (Equipment) contains the core equipment requirements:

211.63 — Equipment Design, Size, and Location

'Equipment used in the manufacture, processing, packing, or holding of a drug product shall be of appropriate design, adequate size, and suitably located to facilitate operations for its intended use and for its cleaning and maintenance.'

Proc-X compliance approach: Equipment design specifications include process-specific sizing calculations. Mixer bowl volumes are validated against minimum fill and maximum fill requirements for the intended batch size range. Equipment location within the facility layout is evaluated during conceptual design for access, maintenance clearance, and cleaning.

211.65 — Equipment Construction

'Equipment shall be constructed so that surfaces that contact components, in-process materials, or drug products shall not be reactive, additive, or absorptive so as to alter the safety, identity, strength, quality, or purity of the drug product beyond the official or other established requirements.'

Proc-X compliance approach: All product-contact surfaces are 316L SS (ASTM A270 standard) unless specific process requirements dictate otherwise (e.g., PEEK or PTFE contact surfaces for highly corrosive API solutions). Material certifications (mill certificates) and surface finish certificates are provided for all product-contact components. No painted surfaces in product zones. No cast iron, zinc, or other reactive metals.

211.67 — Equipment Cleaning and Maintenance

'Equipment and utensils shall be cleaned, maintained, and sanitized at appropriate intervals to prevent malfunctions or contamination that would alter the safety, identity, strength, quality, or purity of the drug product.'

Proc-X compliance approach: Equipment is designed specifically for cleanability. All internal radii ≥ 3 mm (no tight inside corners that trap material). Spray balls and CIP circuits sized for full internal coverage, verified by riboflavin spray tests. Cleaning validation support package included in delivery documentation.

211.68 — Automatic, Mechanical, and Electronic Equipment

'Automatic, mechanical, or electronic equipment or other types of equipment, including computers, or related systems that will perform a function satisfying a requirement of this part shall be calibrated, inspected, or checked according to a written program designed to assure proper performance.'

'Input to and output from the computer or related system of formulas or other records or data shall be checked for accuracy... The degree and frequency of input/output verification shall be based on the complexity and reliability of the computer or related system.'

Proc-X compliance approach: All Proc-X control systems (PLC, SCADA, HMI) are supplied with a full 21 CFR Part 11 compliance package: electronic records with complete audit trail, role-based access control, electronic signature capability, and system validation documentation (IQ/OQ/PQ protocols). See Chapter 30 for detail.

REGULATORY INSIGHT *FDA 483 observations related to equipment are among the most frequent findings in pharmaceutical facility inspections. The top equipment-related 483 categories include: (1) inadequate cleaning validation — using swab and rinse sample recovery data that was never actually demonstrated to correlate with visual inspection endpoints; (2) equipment not of suitable design — crevices, dead legs, and product-retention areas not identified in the equipment qualification; (3) computer systems without adequate audit trails. Proc-X designs against all three proactively.*

3.2 21 CFR Part 111 — Dietary Supplement cGMP

21 CFR Part 111 governs the manufacture, packaging, labeling, and holding of dietary supplements sold in the United States. While less prescriptive than Part 211, Part 111 establishes the same fundamental equipment requirements: proper design for intended use, cleanable construction, and maintenance programs. Key distinctions from Part 211:

- Part 111 requires component identity testing but does not mandate blend uniformity testing unless the manufacturer determines it is necessary — creating a risk area that forward-thinking nutraceutical manufacturers address voluntarily.
- Part 111 requires equipment cleaning SOPs and records but does not require formal cleaning validation studies. However, the expectation of 'adequate cleaning' creates de facto validation requirements for high-risk situations (cross-contamination of allergens, botanicals with active compounds at low label claims).
- 21 CFR Part 111.27 requires that 'Equipment and utensils must be of appropriate design, construction, and workmanship to enable them to be suitable for their

intended use and to be adequately cleaned and properly maintained.' This is nearly identical language to 211.63 and 211.65.

PROC-X EQUIPMENT SPOTLIGHT Proc-X pharmaceutical-grade equipment — designed to meet 21 CFR Part 211 — is inherently compliant with Part 111. Nutraceutical manufacturers who specify Proc-X equipment get pharmaceutical-grade engineering at a price point appropriate for the nutraceutical market, with the added benefit that if they ever choose to manufacture an OTC drug product, their equipment is already qualified to the higher standard.

3.3 EU GMP Annex 11 and ICH Q10

For pharmaceutical manufacturers supplying EU markets, the European Commission's EU GMP Annex 11 (Computerised Systems) governs computer systems used in GMP-regulated environments. Annex 11 requirements for computerized equipment systems include: risk assessment, validation, data integrity controls, audit trails, backup and recovery, and periodic review. The ICH Q10 Pharmaceutical Quality System guideline provides a global framework for quality management that encompasses equipment lifecycle management from design through retirement. Proc-X systems are designed to satisfy both Annex 11 and ICH Q10 requirements.

3.4 How Equipment Design Prevents Regulatory Findings

Regulatory Requirement	Common Finding When Not Met	Proc-X Design Response
211.65 — Non-reactive construction	Metal contamination in product; leachables in dissolution media	316L SS, electropolished; full material certification package
211.67 — Cleanable design	Visual residue after cleaning; positive swab samples	Ra ≤ 0.5 µm, all radii ≥ 3 mm, spray ball coverage mapping
211.68 — Electronic records	No audit trail; shared generic login; data integrity failure	21 CFR Part 11 compliant SCADA, role-based access, electronic signatures
211.68 — Calibration	Calibration records missing or expired	Instrument calibration schedule in DQ package; IQOQPQ protocols
211.63 — Appropriate size	Segregation failure from equipment overload or underfill	Min/max fill validation included in OQ protocol
211.182 — Equipment logs	Equipment log not completed; prior batch cross-contamination	Automatic equipment log generation in PX-BMS batch records

PART II

SOLID DOSAGE FORMS

Granulation, Blending, Compression & Coating

Chapter 4: Granulation Science — Wet, Dry, and Continuous

IN THIS CHAPTER

- ▶ Why granulation exists — the physics of powder flow, compressibility, and content uniformity
- ▶ Wet granulation mechanisms: nucleation, growth, consolidation, and breakage
- ▶ Dry granulation and roller compaction: the solvent-free alternative
- ▶ High-shear vs. fluid bed wet granulation: critical differences
- ▶ Continuous granulation: the process analytical technology (PAT) revolution

4.1 Why Granulation?

Powders are notoriously difficult to process directly into tablets. Fine API powders and excipients typically suffer from poor flowability (which causes tablet weight variation), low bulk density (which requires impractically large tablet presses), poor compressibility (which produces soft tablets that fail hardness and friability specifications), and segregation tendencies (which destroy content uniformity). Granulation solves all of these problems by converting the fine primary powder particles into larger, denser, more uniform granules with controlled properties.

For an API with poor flow and compressibility — which describes the majority of new chemical entities (NCEs) — granulation is not optional. It is the only way to achieve the CQA targets for weight variation ($< 5\%$ RSD) and content uniformity (acceptance value $\leq L1$ per USP $\langle 905 \rangle$) that regulators require.

4.2 Wet Granulation Mechanisms

In wet granulation, a granulating liquid (water, ethanol, or aqueous binder solution) is added to the powder blend, and granules form through a sequence of mechanical and physicochemical events:

- **Nucleation:** Liquid bridges form between primary particles at binder-addition sites. Nucleus granules (50–200 μm) form.
- **Growth:** Nuclei grow by layering (coalescence of smaller granules) and coalescence (collision between nuclei). Granule size distribution broadens.
- **Consolidation:** Granule porosity decreases as particles are packed more tightly under mechanical action.
- **Breakage:** At high impeller speeds or over-wetted conditions, granules fracture back to smaller units. Breakage and growth equilibrate to define the final granule size distribution.

The balance between these mechanisms determines the critical process parameters (CPPs) for the granulation: impeller speed, chopper speed, liquid addition rate, wet

massing time, and binder concentration. Proc-X high-shear granulators provide independent control of all these variables with data logging for every batch.

ENGINEERING INSIGHT *The granule size distribution is the primary process output of wet granulation, and it directly determines tablet properties downstream. Granules too fine ($D_{50} < 150 \mu\text{m}$): poor flow, tablet weight variation, segregation during blending. Granules too coarse ($D_{50} > 1,000 \mu\text{m}$): poor compressibility, speckled appearance, slow dissolution. The CPP-to-CQA relationship for granule size is the core design space element in any QbD-based granulation development program.*

4.3 Dry Granulation and Roller Compaction

Dry granulation avoids liquid entirely — critical when the API or binder is moisture-sensitive, or when wet granulation drying times are prohibitively long. The roller compactor presses powder between two counter-rotating rolls at 3–50 kN/cm line force, producing a dense ribbon or flake that is then milled (typically in an oscillating granulator or Comil) to the target granule size.

Critical roller compaction parameters and their effects:

Parameter	Range	Effect on Ribbon / Granule	Measurement
Roll pressure	3–50 kN/cm	Higher pressure → denser ribbon, harder granule, slower dissolution	Inline force transducer
Roll speed	1–10 rpm	Higher speed → less dwell time, softer ribbon if pressure constant	Encoder feedback
Feed screw speed	5–60 rpm	Controls powder feed rate and compaction zone fill	Motor current + torque
Ribbon relative density	0.65–0.85 g/cc	Primary CQA for roller compaction; predictor of granule flow	NIR or inline density
Milling screen size	0.5–3.0 mm	Controls granule D90 from ribbon milling step	Sieve analysis, Mastersizer

PROC-X EQUIPMENT SPOTLIGHT Proc-X PX-RC roller compactors feature a patented gap control system that maintains constant roll gap ($\pm 0.01 \text{ mm}$) regardless of feed rate variation, eliminating the ribbon density variation caused by powder flow fluctuations in conventional force-fed compactors. An inline NIR probe on the ribbon discharge monitors ribbon relative density in real time, with automatic roll force adjustment to maintain the target — achieving $Cpk \geq 1.67$ for ribbon density at steady state.

4.4 Continuous Granulation

The pharmaceutical industry's shift toward continuous manufacturing (CM) has brought twin-screw wet granulation (TSG) to the forefront. In TSG, powder and granulating liquid are continuously fed into a co-rotating twin-screw extruder at controlled rates, granulation occurs within the screw barrel in 20–60 seconds, and granules are discharged continuously to a downstream fluid bed dryer. The entire process is monitored with PAT (Process Analytical Technology) instruments:

- Inline particle size: Focused Beam Reflectance Measurement (FBRM) or spatial filter velocimetry (SFV) measures granule size in real time in the transfer line from granulator to dryer.

- Inline moisture: NIR spectroscopy measures granule moisture content in the fluid bed dryer discharge, replacing offline LOD (loss-on-drying) measurements.
- Inline blend uniformity: NIR or Raman spectroscopy measures API content in the granule stream, enabling real-time release testing (RTRT).

Chapter 5: High-Shear Wet Granulation Equipment

IN THIS CHAPTER

- ▶ High-shear granulator design: bowl, impeller, chopper, and their functions
- ▶ Torque and power monitoring as primary process endpoints
- ▶ Scale-up from laboratory to production: the impeller tip speed principle
- ▶ Proc-X PX-HSG series: design features, bowl sizes, and cGMP compliance
- ▶ In-process controls: water addition endpoint, wet massing endpoint, granule sampling

5.1 Equipment Architecture

The high-shear granulator (HSG) consists of a hemispherical or cylindrical mixing bowl, a bottom-driven impeller rotating at 100–600 rpm, and a side-mounted chopper rotating at 1,000–3,600 rpm. The impeller generates the bulk mixing action and supplies the mechanical energy for granule growth and consolidation. The chopper breaks up lumps and oversize agglomerates formed during wet massing, controlling the upper end of the granule size distribution.

Component	Design Variable	Effect on Granulation	Proc-X Specification
Impeller	Blade angle (45°–90°)	Steeper angle → more radial flow, faster liquid distribution	Three-blade, 45° standard; custom geometries available
Impeller	Tip speed (2–12 m/s)	Higher tip speed → smaller, denser granules	VFD control, ± 0.1 m/s; logged every second
Chopper	Speed (1,000–3,600 rpm)	Higher speed → narrower size distribution	Independent VFD; explosion-proof motor option
Bowl	Volume (2–600 L)	Determines batch size range	2, 10, 25, 65, 150, 300, 600 L standard
Bowl jacket	Temperature (5–60°C)	Controls product temperature during granulation	Glycol jacket, $\pm 1^\circ\text{C}$; dual thermocouple monitoring
Spray nozzle	Droplet size, spray rate	Larger droplets → overwetting risk; slower spray → longer process	Binary nozzle, 50–500 mL/min per nozzle, peristaltic control

5.2 Power and Torque as Process Endpoints

The most reliable process endpoint for high-shear wet granulation is motor power (or torque) monitoring. As binder liquid is added, the powder blend becomes a wet mass, and motor power increases. The power profile — specifically the rate of rise and the peak — correlates with granule size and density and is used to define the end of wet massing:

- Level 1 (power onset): Powder bed has been wetted; binder is distributed. Granules are forming.

- Level 2 (power plateau): Granule growth has equilibrated; mean granule size is approaching target.
- Level 3 (power peak / rapid rise): Over-wetting risk. Batch must be discharged immediately or liquid addition must stop.

ENGINEERING INSIGHT *A frequently underappreciated source of batch-to-batch variability in high-shear granulation is the humidity and temperature of the granulation room. On a humid day, the same binder liquid addition can produce a wetter wet mass and smaller, denser granules than on a dry day — because the powder has already absorbed atmospheric moisture before granulation begins. Proc-X granulators with HVAC-interlocked jacketed lids and real-time humidity monitoring inside the bowl address this variability source proactively.*

5.3 Scale-Up Principle: Impeller Tip Speed

The most validated scale-up rule for high-shear granulation is constant impeller tip speed: $v_{tip} = \pi \times d \times n / 60$. If v_{tip} is held constant from lab scale to production scale, and the liquid-to-solid ratio (L/S ratio) is unchanged, the granule size distribution is generally preserved across scales. Proc-X documents the tip speed history for every batch, making scale-up validation a straightforward comparison of tip speed profiles across scales.

PROC-X EQUIPMENT SPOTLIGHT Proc-X PX-HSG series high-shear granulators are available in 2, 10, 25, 65, 150, 300, and 600 L bowl sizes with matched impeller geometry scaled by constant tip speed. All models: top-driven (no bottom bearing exposed to product), cGMP polished 316L SS bowl, integrated spray nozzle system (binary atomizing nozzle, peristaltic pump, mass flow meter), real-time power and torque monitoring logged to PX-BMS, and vacuum discharge to contained dryer transfer. OEB 4 contained charging available for all sizes.

Chapter 6: Fluid Bed Processing — Granulation, Drying & Coating

IN THIS CHAPTER

- ▶ Fluid bed physics: how a fluidized powder bed enables three distinct unit operations
- ▶ Top-spray granulation: mechanism, binder distribution, and key parameters
- ▶ Bottom-spray (Wurster) coating: sustained/delayed/enteric release coating
- ▶ Tangential-spray (rotor) processing for pellet coating and densification
- ▶ Spray drying integration and the fluid bed as a continuous dryer

6.1 The Fluid Bed as a Multi-Function Platform

The fluid bed processor is the most versatile single piece of equipment in solid dosage form manufacturing. In fluid bed operation, heated process air is forced upward through a perforated distribution plate at sufficient velocity to suspend the powder or granule bed in a fluidized state. In this state, every particle is in continuous motion and fully exposed to the surrounding air stream — enabling rapid heat and mass transfer. This single characteristic enables three distinct unit operations:

- Drying: Water (or solvent) evaporation from wet granules after high-shear or tumble granulation. Fluid bed drying is 3–5× faster than tray drying for equivalent moisture removal.
- Top-spray granulation: Binder solution sprayed down onto the fluidized powder bed from nozzles above. Wetting, agglomeration, drying — all in a single process step.
- Coating: Film-forming coating solution sprayed onto the surface of granules, pellets, or tablets from below (Wurster insert) or tangentially (rotor insert) to produce immediate release, delayed release, enteric, or sustained release film coatings.

6.2 Top-Spray Granulation Process Parameters

Parameter	Range	CQA Impact	Control
Inlet air temp	40–80°C	Granule moisture, granule density	PID temperature controller
Process air volume	200–2,000 m ³ /hr (scale-dependent)	Fluidization quality, spray drying vs. granulation balance	VFD fan speed + differential pressure control
Spray rate	5–200 g/min per nozzle	Granule size: slow rate → larger granules; fast → finer	Peristaltic pump, gravimetric control
Nozzle air pressure	0.5–3.0 bar	Droplet size: higher pressure → smaller droplets → finer granules	Precision pressure regulator
Product temperature	30–55°C	Primary process endpoint: correlates with moisture	Inline thermocouple, NIR probe

Binder concentration	2–20% w/w	Higher conc → larger granules, more friable granule bed	Binder solution prep + gravimetric verification
----------------------	-----------	---	---

6.3 Wurster (Bottom-Spray) Coating

The Wurster coating insert produces the most uniform film coatings achievable in pharmaceutical manufacturing. A central partition tube creates a high-velocity air jet that lifts individual particles up through the coating zone, where they are exposed to the spray nozzle mounted below. Particles exit the top of the Wurster tube, travel outward through the expansion chamber (where the film dries), and re-enter the bottom of the Wurster column for the next coating pass. Each particle makes thousands of coating passes, and the film builds uniformly with each pass.

Applications: enteric coating (pH-responsive polymers — Eudragit L30 D-55, HPMCAS), sustained release (ethylcellulose, Surelease), taste masking (HPMC, Opadry), delayed release (HPMCP). Film weight gain controls the release profile, and uniformity of film weight across the bed is the critical CQA.

PROC-X EQUIPMENT SPOTLIGHT Proc-X PX-FBP fluid bed processors are available in 5, 15, 30, 60, 120, 200, and 400 kg batch sizes with modular product container design. Each system ships with top-spray, Wurster, and rotor inserts included — no additional capital cost for process conversion between granulation and coating operations. Automated spray rate ramping, product temperature feedback control, and inline LOD measurement are standard on all models.

Chapter 7: Powder Blending for Solid Dosage Manufacturing

IN THIS CHAPTER

- ▶ Blending mechanisms: convective, diffusive, shear — and when each dominates
- ▶ V-blenders: the workhorse of pharmaceutical blending
- ▶ Bin blenders: the contained blending solution for OEB 3–5 APIs
- ▶ Blend endpoint determination: thief sampling vs. PAT
- ▶ Segregation: causes, consequences, and equipment solutions

7.1 Blending Mechanisms and Equipment Selection

Pharmaceutical blending involves three fundamental mechanisms. Convective mixing: bulk movement of powder groups from one region to another — dominant in low-speed tumble blenders and ribbon blenders. Diffusive mixing: random movement of individual particles across interfaces — dominant in high-speed blenders. Shear mixing: slip planes developing between layers of different concentrations. The relative contribution of each mechanism, and the optimal equipment choice, depends on the particle properties (size, density, shape, cohesion) and the target blend uniformity.

Blender Type	Mechanism	Typical Batch	Key Advantage	Key Limitation
V-blender	Convective + diffusive	5–2,000 kg	Simple, cleanable, scalable	Slow for cohesive powders; limited for OEB 4+
Bin blender (IBC)	Convective + diffusive	50–1,500 kg	Contained transfer; no sieving cross-contamination	Requires dedicated IBC investment
Ribbon blender	Convective + shear	20–5,000 kg	Fast for large batches; handles wet masses	Difficult to clean; segregation on discharge
High-shear mixer (blending mode)	Convective + dispersive	2–600 kg	Fast blend for cohesive powders	Heat generation at high speeds
Tumble blender (octagonal)	Convective + diffusive	5–500 kg	Gentle for fragile granules/pellets	Slow; limited shear for deagglomeration

7.2 Bin Blenders for OEB 3–5 Processing

The bin blender (also called IBC — Intermediate Bulk Container — blender) has become the preferred blending solution for potent compounds. The IBC itself is both the blending vessel and the transfer container — powder is charged into the IBC at the dispensing step (via a contained charging isolator for OEB 4/5), the IBC is transported to the blender, blended by tumbling on the bin blender drive, then transported directly to the tablet press or capsule filler and discharged without further transfer. The API never contacts room air after the contained charging step.

PROC-X EQUIPMENT SPOTLIGHT Proc-X PX-IBC bin blenders are available in 50 to 1,500 L IBC sizes and drive capacities up to 2,000 kg total loaded weight. The drive uses a horizontal roll-over tumbling action at 6–20 rpm with independent rotation speed control. All PX-IBC models include: automatic IBC recognition (RFID coupling), weight verification before and after blend (integral load cells), and PX-BMS data logging of every rotation counted. Contained charging stations (OEB 4: < 1 µg/m³; OEB 5: < 0.1 µg/m³) supplied as matched systems.

7.3 Blend Endpoint Determination

The traditional method for blend endpoint determination — thief sampling followed by HPLC or UV analysis — is slow (24–48 hour turnaround), invasive (sampling disturbs the blend), and subject to thief sampling error (particle segregation during sampling). The modern approach uses PAT:

- NIR spectroscopy: NIR probes mounted in the blender wall or IBC chute monitor API spectral signature in real time. When blend uniformity (RSD < 3.0%) is achieved, the blend is considered complete — regardless of elapsed time. This eliminates both over-blending (which can cause segregation by reducing particle cohesion) and under-blending.
- Moving block standard deviation (MBSD): A statistical analysis of the rolling RSD of NIR signals over a defined time window detects when the blend has reached a steady, uniform state.

Chapter 8: Film Coating and Pan Coating Systems

IN THIS CHAPTER

- ▶ Coating objectives: immediate release, extended release, enteric, taste masking
- ▶ Side-vented coating pan design and its advantages over conventional coating pans
- ▶ Spray system design: atomization, spray pattern, and coating uniformity
- ▶ Process parameters: pan speed, inlet temperature, spray rate, and air flow
- ▶ Coating defects and their root causes in processing

8.1 Why Coat Tablets?

Tablet coating serves multiple functions: taste masking (APIs with bitter taste), moisture protection (hygroscopic APIs), light protection, immediate release (smooth swallowability), extended release (rate-controlling polymer membrane), and enteric release (pH-responsive polymer preventing gastric dissolution). Each functional coating type has its own polymer system, process requirements, and critical quality attributes.

Coating Type	Polymer System	Key CQA	Process Challenge
Immediate release (IR)	HPMC (Opadry), PVA	Film uniformity, disintegration time	Spray rate balance vs. drying capacity
Enteric	HPMCAS, Eudragit L100-55	Acid resistance (< 10% dissolution in 0.1N HCl × 2h)	Coating weight uniformity; sticky spray
Sustained release (SR)	Ethylcellulose, Aquacoat	In vitro dissolution profile match	Film uniformity; permeability control
Extended release (ER)	Eudragit RS/RL blend	Release rate and variability	Blend ratio precision; process temp control
Taste masking	Eudragit E100, HPMC	Dissolution in saliva (< 60 sec); taste panel	Thin, uniform coat; no film bridging

8.2 Side-Vented Coating Pan

The side-vented perforated coating pan is the standard production coating system in pharmaceutical manufacturing. Tablets tumble in a rotating drum; coating solution is sprayed by spray guns mounted on a spray arm inside the drum; process air flows through perforations in the drum wall for drying. Side-vented design allows continuous air flow through the tablet bed, achieving much faster drying (and thus higher spray rates) than conventional solid-wall pans.

Key design parameters: drum diameter (300 mm laboratory to 2,000 mm production), fill volume (40–75% of drum volume), drum speed (3–20 rpm), and baffling configuration (straight baffles, zig-zag baffles, or no baffles for fragile tablets).

PROC-X EQUIPMENT SPOTLIGHT Proc-X PX-COAT side-vented coating pans range from 10 kg laboratory scale to 600 kg production scale. All include: infrared exhaust temperature measurement for process endpoint control, automated spray rate ramping protocol (prevents initial tablet wetting), and inlet/exhaust air dew point monitoring. The PX-COAT integrated PAT interface accepts inline NIR or UV probes for film weight gain monitoring without manual sampling.

Chapter 9: Twin-Screw Extrusion and Hot-Melt Extrusion in Pharma

IN THIS CHAPTER

- ▶ Hot-melt extrusion (HME): the most powerful tool for bioavailability enhancement
- ▶ Amorphous solid dispersions: why they work and how HME makes them
- ▶ TSE screw design for pharmaceutical applications
- ▶ Downstream processing of HME extrudate: milling and tableting
- ▶ Continuous wet granulation via twin-screw extrusion

9.1 Hot-Melt Extrusion and Bioavailability Enhancement

Approximately 40% of approved drug products and 70% of development pipeline compounds are classified as BCS Class II (high permeability, low solubility) — meaning their bioavailability is dissolution-rate limited. Hot-melt extrusion (HME) is the most effective approach to solving this problem for thermally stable APIs. In HME, the API is dispersed (or dissolved) in a polymer matrix at elevated temperature, producing an amorphous solid dispersion (ASD) with dramatically improved dissolution kinetics.

The science: crystalline APIs have ordered lattice structures with high cohesive energy — they dissolve slowly. An amorphous form of the same API has no lattice — it is a supercooled liquid. Amorphous APIs dissolve 10–1,000× faster than their crystalline forms. An ASD produced by HME stabilizes the amorphous form by trapping the API in a polymer matrix (typically HPMC-AS, PVP-VA, or Eudragit EPO) that prevents recrystallization, both during storage and during dissolution in GI fluid. The result: supersaturated API solutions in the GI tract, far exceeding the crystalline solubility, and dramatically improved bioavailability.

9.2 TSE Screw Design for HME

- Conveying elements: move material forward and build pressure. Located at the feed zone and in the die approach zone.
- Kneading blocks: provide distributive and dispersive mixing, melt the polymer, and distribute API in the melt. Most critical elements for ASD quality — block stagger angle (30°, 45°, 60°, 90°) and width determine shear intensity.
- Mixing elements: additional mixing after kneading. Used when API–polymer miscibility requires high shear.
- Reverse conveying elements: increase residence time and back-pressure in the mixing zone for APIs needing extended mixing for dissolution in the polymer melt.

HME Parameter	Range	Effect on ASD	Measurement
Barrel temperature profile	80–200°C (zone-dependent)	Above Tg of polymer required; below API degradation temp	Thermocouple each zone; API stability limit as hard stop

Screw speed	50–600 rpm	Higher speed → more shear, faster melt, shorter residence time	Encoder; melt temperature inline
Throughput rate	0.1–50 kg/hr	Affects residence time and melt temperature at fixed screw speed	Gravimetric feeder accuracy $\pm 0.1\%$
Die pressure	200–1,500 psi	Affects extrudate density and shape consistency	Inline pressure transducer
Melt temperature	100–230°C	If too high: API degradation; too low: incomplete dissolution in polymer	Inline melt thermocouple at die
Residence time distribution	30–180 seconds	Narrow RTD → uniform ASD; wide RTD → amorphous content variation	Tracer dye experiment (OQ)

PROC-X EQUIPMENT SPOTLIGHT Proc-X PX-HME twin-screw extruders for pharmaceutical HME are available in 11 mm (laboratory, 0.1–1 kg/hr), 24 mm (development, 0.5–5 kg/hr), and 40 mm (production, 5–50 kg/hr) screw diameters. All models feature: barrel temperature uniformity $\pm 1^\circ\text{C}$, gravimetric loss-in-weight feeders for API and polymer, inline melt pressure and temperature monitoring, co-rotating screw design for superior mixing, and PX-BMS 21 CFR Part 11 compliant data logging. GMP-grade stainless steel contact surfaces with $R_a \leq 0.8 \mu\text{m}$.

PART III

API PROCESSING

Milling, Micronization & Nanosizing

Chapter 10: The API Particle Size Challenge

IN THIS CHAPTER

- ▶ Why API particle size is a critical quality attribute in almost every dosage form
- ▶ The Noyes-Whitney equation: the physical chemistry of dissolution and its processing implications
- ▶ Particle size specifications in pharmacopoeias: USP <429>, USP <811>, ICH Q6A
- ▶ Polymorphism, crystal habit, and surface area — and how milling affects all three
- ▶ Occupational exposure and containment requirements for API milling

10.1 Why Particle Size Matters

Particle size is a critical quality attribute (CQA) for APIs in virtually every dosage form category. The physical reasons are grounded in fundamental pharmacokinetics and dosage form physics:

- **Dissolution rate:** For BCS Class II APIs, the Noyes-Whitney equation governs dissolution rate: $dM/dt = (D \times A \times C_s \times [C_s - C_b]) / h$. Reducing particle size increases surface area A , which directly increases dM/dt . A 10-fold reduction in particle diameter produces approximately a 100-fold increase in specific surface area — and correspondingly faster dissolution and absorption.
- **Content uniformity:** For low-dose tablets ($API \leq 5\%$ w/w), particle size of the API directly determines the statistical probability of achieving the required content uniformity. USP <905> uses a sampling theory model that shows that for a 5 mg dose in a 200 mg tablet, a D90 of 125 μm is acceptable, but a D90 of 250 μm may cause unacceptable content uniformity variability.
- **Aerosol performance:** Dry powder inhalers (DPIs) require API particles with mass median aerodynamic diameter (MMAD) of 1–5 μm to achieve adequate deep lung deposition. This requires jet milling to the micron range, with extremely tight particle size specifications (MMAD $2.5 \pm 0.5 \mu m$ in some formulations).
- **Physical and chemical stability:** Smaller API particles have higher surface energy and are more susceptible to polymorphic conversion, hygroscopic uptake, and oxidative degradation. The optimal particle size balances dissolution performance against stability risk.

10.2 The Noyes-Whitney Equation in Practice

$dM/dt = (D \times A \times C_s) / h$ (simplified for sink conditions, where $C_b \ll C_s$)

D = diffusion coefficient in GI fluid ($\sim 5 \times 10^{-6} \text{ cm}^2/\text{s}$ for most small-molecule drugs), A = surface area of undissolved particles, C_s = intrinsic solubility, h = diffusion layer thickness. Of these parameters, A is the one most directly controlled by processing. For a given mass

of API, $A \propto 1/d$, where d is particle diameter. Reducing d from 100 μm to 1 μm increases A by 100-fold and increases dM/dt by 100-fold.

REGULATORY CONTEXT *The European Pharmacopoeia and USP define particle size by equivalent volume diameter using laser diffraction (Ph. Eur. 2.9.31, USP <429>). The specification format is typically: $D(0.1) \geq X \mu\text{m}$, $D(0.5) = Y \mu\text{m} \pm Z\%$, $D(0.9) \leq W \mu\text{m}$. For inhalation products, aerodynamic particle size distribution (APSD) by cascade impactor (Ph. Eur. 2.9.18, USP <601>) is the required method. Proc-X jet mill systems include a factory-tested PSD characterization using laser diffraction as part of the OQ protocol.*

10.3 Polymorphism and Milling

Milling can induce polymorphic transitions in APIs — a critical concern when the active polymorph has different dissolution or stability properties than the inactive form. The high local temperature and shear stress generated by impact milling (hammer mills, pin mills) can cause surface amorphization or polymorph conversion. Jet milling with temperature-controlled nitrogen carrier gas avoids this by keeping the milling zone at a controlled, low temperature. Cryogenic jet milling (with liquid nitrogen cooling to $< -50^\circ\text{C}$) provides maximum protection against milling-induced polymorphism for thermally sensitive APIs.

Chapter 11: Jet Milling and Micronization of APIs

IN THIS CHAPTER

- ▶ How the opposed jet mill produces micronized API with a tight particle size distribution
- ▶ Critical process parameters and their effect on API PSD
- ▶ Nitrogen carrier gas systems for API milling: why air is usually not acceptable
- ▶ cGMP design requirements for pharmaceutical jet mills
- ▶ Scale-up from laboratory to production scale for API jet milling

11.1 Jet Milling Mechanism

In an opposed jet mill, two high-velocity gas jets (nitrogen or air at 80–150 psig) are directed at each other. API particles fed into the mill are accelerated by the gas jets and collide in the inter-jet zone at velocities of 300–500 m/s. Particle–particle collisions provide the energy for size reduction. A classification wheel (rotor) at the center of the mill rotates at 2,000–15,000 rpm, creating a centrifugal field that returns oversize particles to the grinding zone while allowing product-specification particles to pass through to the collector.

The opposed jet mill's key advantages for API processing: no mechanical grinding surfaces in contact with the product (eliminating metal contamination risk), no heat generation at surfaces (protecting thermolabile APIs), extremely tight particle size distributions achievable (D10/D90 span < 3), and full containment capability for high-potency APIs.

Parameter	Range	Effect on API PSD	cGMP Monitoring
Classifier wheel speed	2,000–15,000 rpm	Primary D50 control — higher speed → finer product	VFD tachometer, logged real-time
Grinding gas pressure	80–150 psig N ₂	Higher pressure → more energy, finer D90	Calibrated transducer, logged
Feed rate	0.1–50 kg/hr	Higher feed → product coarsening if classifier not adjusted	Gravimetric feeder, ± 0.5%
Nitrogen gas temperature	–20°C to +25°C	Cold N ₂ protects thermolabile APIs; prevents sticking	Inline thermocouple
Classifier cut point (D50 target)	1–30 µm (API typical)	Determines primary CQA — dissolution performance	Inline laser diffraction preferred

11.2 Nitrogen Gas Systems for API Milling

Pharmaceutical API milling almost universally requires nitrogen as the carrier and grinding gas rather than compressed air. Reasons:

- Oxidation prevention: Many APIs are oxidation-sensitive. Nitrogen ($O_2 < 1\%$) prevents API surface oxidation that occurs within seconds on fine, high-surface-area particles in air.
- Moisture exclusion: Compressed air, even after drying, contains trace moisture that can affect hygroscopic APIs and cause product agglomeration in the classifier.
- Explosion prevention: Fine API powders at the concentrations present in a jet mill can be in the explosive dust range. Nitrogen inerting prevents the oxygen atmosphere needed for dust explosion.

PROC-X EQUIPMENT SPOTLIGHT Proc-X PX-API jet mill systems are supplied as complete, validated pharmaceutical milling systems: the mill, N₂ supply and control, inline O₂ monitoring (with alarm and auto-shutoff at $> 2\%$ O₂), product collection (contained cyclone + HEPA filter system), and contained discharge under N₂ blanket. All product-contact surfaces: 316L SS, electropolished to $Ra \leq 0.5 \mu m$. Available with OEB 4 ($< 1 \mu g/m^3$) and OEB 5 ($< 0.1 \mu g/m^3$) containment configurations for high-potency APIs.

11.3 Scale-Up for API Jet Milling

Scale-up from laboratory jet mill to production scale is governed by the principle of constant classifier tip speed ($v_{tip} = \pi \times d_{classifier} \times n / 60$). If the classifier tip speed is held constant from lab to production, and the specific energy input per kilogram of feed is maintained, the PSD typically scales with accuracy $\pm 10\%$ on D₅₀. Proc-X provides laboratory systems (100 g/hr to 2 kg/hr), pilot systems (2–20 kg/hr), and production systems (20–500 kg/hr) with documented scale-up packages for each platform.

Chapter 12: Wet Milling, Nanosizing, and Media Milling

IN THIS CHAPTER

- ▶ When wet milling is better than jet milling for API size reduction
- ▶ Nanosizing: the science, the regulatory pathway, and why $< 1\ \mu\text{m}$ sometimes matters
- ▶ Planetary ball mills and bead mills for wet API nanosizing
- ▶ Stabilizer selection and suspension formulation for nanosized APIs
- ▶ Scale-up challenges unique to wet API milling

12.1 Wet Milling vs. Jet Milling

Wet milling suspends the API in a liquid medium (typically water with a stabilizer/surfactant) and mills using a planetary ball mill or bead mill. It is preferred over jet milling when:

- Target particle size $< 1\ \mu\text{m}$: Jet mills reliably achieve $D_{50} \geq 1\ \mu\text{m}$. Wet bead mills achieve D_{50} down to 100–200 nm consistently.
- API has very low density: Low-density APIs accelerate poorly in a gas jet, limiting micronization efficiency. Wet milling is size-reduction mechanism independent of particle density.
- API is thermolabile: Even cold-gas jet mills introduce brief surface temperature spikes. Wet milling with jacketed systems can maintain $< 10^{\circ}\text{C}$ throughout.
- Liquid dispersion is the final form: For parenteral suspensions, lyophilized injectables, or oral suspensions, wet milling produces the final suspension directly, eliminating a redispersion step.

12.2 Nanosizing Technology

Nanosized API particles (100–1,000 nm) represent the ultimate solution to dissolution-limited bioavailability for BCS Class II compounds. A 200 nm particle has 500× the specific surface area of a 100 μm particle. When dosed orally as a nanoparticulate suspension or as a dry redispersible formulation (produced by spray drying a nanosuspension), nanosized APIs achieve AUC values approaching IV administration for some compounds.

Wet bead milling with 0.1–0.3 mm zirconia or polystyrene beads at 2,000–4,000 rpm achieves nanosized API suspensions in 4–24 hours (batch) or 30–120 minutes per pass (continuous). Stabilizers (Poloxamer 188, HPMC, SDS, PVP) are essential to prevent Ostwald ripening (particle growth) and maintain the nanosuspension during downstream processing.

ENGINEERING INSIGHT *The most common failure mode in API nanosizing is Ostwald ripening during storage — the thermodynamic process by which small particles dissolve and redeposit on larger particles, increasing mean particle size over time. The solution is proper stabilizer selection: Poloxamer 188 at 0.5–2% is effective for hydrophobic BCS*

Class II APIs; HPMC-AS provides both steric stabilization and precipitation inhibition during dissolution. The selection of stabilizer type and concentration should be part of the nanosizing formulation development program, not an afterthought.

Chapter 13: Cryogenic Milling for Thermolabile APIs

IN THIS CHAPTER

- ▶ Why ambient milling fails for thermolabile and thermoplastic APIs
- ▶ Cryogenic jet milling and cryogenic pin milling systems
- ▶ Liquid nitrogen usage, safety, and facility requirements
- ▶ APIs requiring cryogenic milling: examples and process parameters

13.1 The Thermolabile Milling Challenge

Many pharmaceutical APIs have low melting points ($< 80^{\circ}\text{C}$), decompose at moderate temperatures, or undergo polymorphic conversion in response to the brief thermal spikes generated during particle–particle collisions in conventional milling. For these APIs, standard jet milling (even with nitrogen) can produce: polymorphic conversion that changes dissolution rate, localized surface decomposition that generates genotoxic impurities, or particle agglomeration caused by surface softening.

Cryogenic milling uses liquid nitrogen (LN2) or liquid CO2 to cool the API feed, the mill interior, and the carrier gas to temperatures of -50°C to -150°C . At these temperatures, APIs below their glass transition temperature (T_g) become brittle glass-like solids that fracture cleanly without plastic deformation, polymer softening, or thermal degradation.

API Category	Milling Challenge	Cryogenic Approach	Target Conditions
Thermoplastic polymorphic APIs	Polymorph conversion at $> 40^{\circ}\text{C}$	LN2 jacketed jet mill, pre-cooled feed	Milling zone $< -30^{\circ}\text{C}$
Low melting point APIs ($T_m < 70^{\circ}\text{C}$)	Particle fusion/agglomeration	Cryogenic pin mill or impact mill	Feed and mill at -50°C to -70°C
Amorphous ASDs (HME extrudate)	Recrystallization on milling energy input	Cryogenic cutter/granulator for extrudate	$< T_g - 30^{\circ}\text{C}$ during size reduction
Protein and peptide APIs	Denaturation, aggregation from mechanical stress	Cryogenic ball mill (precooled, gentle)	-196°C (LN2 bath), very low energy
Moisture-sensitive APIs	Hygroscopic uptake $\rightarrow$ agglomeration	Cryogenic milling + N2 dry gas sweep	Dew point $< -40^{\circ}\text{C}$ in milling zone

PROC-X EQUIPMENT SPOTLIGHT Proc-X cryogenic milling systems use a proprietary pre-cooling feed screw that reduces API temperature from ambient to below T_g before the feed enters the grinding chamber, eliminating the transient thermal spikes that occur when ambient-temperature feed contacts the cold grinding chamber walls. LN2 consumption is reduced by 35% versus competing systems that rely entirely on gaseous N2 purge cooling of the mill chamber.

Chapter 14: Continuous API Manufacturing and PAT Integration

IN THIS CHAPTER

- ▶ ICH Q13 Continuous Manufacturing — the regulatory framework that makes CM viable
- ▶ The continuous API processing line: from synthesis to finished API powder
- ▶ PAT instruments for continuous API processing: NIR, Raman, FBRM, laser diffraction
- ▶ Real-time release testing (RTRT): what it requires from processing equipment

14.1 ICH Q13 and Continuous Manufacturing

ICH Q13 (Continuous Manufacturing of Drug Substances and Drug Products), finalized in 2022, established the global regulatory framework for continuous pharmaceutical manufacturing. Key principles from ICH Q13 directly relevant to processing equipment:

- Control strategy: The control strategy for CM must address system dynamics — how the process responds to disturbances, and how monitoring and control systems detect and respond to out-of-range conditions.
- Residence time distribution (RTD): CM processes must characterize and control the RTD — the distribution of time that material spends in the process. A wide RTD means material from different input lots can exit simultaneously, complicating batch definition.
- Diversion strategy: Procedure for diverting potentially out-of-spec material must be established. Proc-X diversion valves with < 0.5-second response time integrated into the control architecture are the standard solution.
- Batch definition in CM: ICH Q13 allows batch definition by time duration, material volume, or material mass — flexible, but requiring clear documentation in the Process Control Description (PCD).

14.2 PAT in Continuous API Processing

PAT instruments provide the real-time quality monitoring that makes continuous manufacturing viable. For continuous API milling and granulation:

- NIR spectroscopy: Identifies API polymorphic form, measures moisture content, and monitors blend uniformity in real time. Sample interval: 1–10 seconds.
- Raman spectroscopy: Higher chemical specificity than NIR. Distinguishes polymorphs with very similar NIR spectra. Also monitors API concentration in wet milling suspensions.
- Inline laser diffraction: Measures particle size distribution of the milled product stream in real time. Classifier wheel speed adjusts automatically when PSD drifts out of specification.

- FBRM (Focused Beam Reflectance Measurement): Chord length distribution measurement in wet granulation and wet milling systems. Provides real-time granule size trend without the sampling errors of offline measurements.

PROC-X EQUIPMENT SPOTLIGHT Proc-X continuous API processing lines include the PX-PAT integration platform — an open-architecture data hub that accepts data streams from any PAT instrument vendor (Metrohm, Mettler-Toledo, Kaiser Optical, Malvern Panalytical, Innopharma) through standard OPC-UA connections. The platform applies multivariate statistical models in real time to detect out-of-spec conditions and trigger automated diversion — no manual intervention required for routine control actions.

PART IV

NUTRACEUTICALS & SUPPLEMENTS

Botanical Extracts, Spray Drying & Encapsulation

Chapter 15: The Nutraceutical Processing Landscape

IN THIS CHAPTER

- ▶ Market structure of the global nutraceutical industry
- ▶ Key raw material categories: botanicals, vitamins, minerals, omega fatty acids, probiotics
- ▶ 21 CFR Part 111 requirements and how they shape processing choices
- ▶ The quality challenge: identity, purity, and potency in botanical processing

15.1 The Nutraceutical Market and Its Processing Demands

The global dietary supplement and nutraceutical market is one of the fastest-growing sectors in consumer healthcare, driven by self-care trends, aging demographics, and growing consumer demand for preventive health products. Behind every capsule of turmeric extract, every protein powder, every omega-3 softgel, and every probiotic tablet is a processing chain that must deliver consistent identity, purity, and potency — now increasingly under the scrutiny of FDA GMP inspections, third-party certifications (NSF, USP-V, Informed Sport), and private-label retailer specifications.

The raw materials in nutraceutical manufacturing span an extraordinary range of physicochemical properties:

- Botanicals: Herbal extracts (spray-dried or liquid), powdered plant materials, standardized extracts. High lot-to-lot variability in marker compound content, moisture, and particle size. Processing challenge: standardization of concentration and particle size.
- Vitamins and minerals: Ascorbic acid, tocopherols, zinc oxide, magnesium glycinate. Highly variable bulk densities and flow properties. Processing challenge: content uniformity in low-dose formulations.
- Omega fatty acids: Fish oil, algal DHA, flaxseed oil. Oil-based, susceptible to oxidation. Processing challenge: encapsulation, oxidative stability, moisture control.
- Probiotics: Live bacterial cultures at 10^8 – 10^{11} CFU/g. Heat-sensitive, moisture-sensitive, oxygen-sensitive. Processing challenge: survival through processing, packaging, and shelf life.
- Sports nutrition: Protein concentrates, amino acids, creatine, beta-alanine. Large-volume, commodity-type ingredients. Processing challenge: blend uniformity, flavor masking, clumping prevention.

15.2 21 CFR Part 111 Processing Requirements

Under 21 CFR Part 111, dietary supplement manufacturers must establish and follow written procedures for all manufacturing operations, maintain equipment cleaning and

maintenance logs, verify component identity before use, and conduct in-process checks for physical characteristics (blend uniformity, moisture, tablet weight, dissolution or disintegration where applicable). Proc-X equipment for nutraceutical manufacturing includes:

- Integrated moisture monitoring for botanical powders and granules — inline NIR reduces the drying time uncertainty that leads to out-of-spec moisture in botanical products.
- Weight verification systems for bin blenders — automatic logging of IBC weight before and after blending provides documentary evidence of batch integrity.
- Clean-in-place (CIP) systems validated to eliminate cross-contamination risk in allergen-managed facilities.

Chapter 16: Spray Drying for Nutraceutical Powder Production

IN THIS CHAPTER

- ▶ Spray drying fundamentals: how a liquid becomes a powder in < 30 seconds
- ▶ Microencapsulation by spray drying: protecting sensitive ingredients
- ▶ Process parameters and their effect on particle properties
- ▶ Spray drying of probiotics, omega oils, and botanical extracts
- ▶ Agglomeration and instant powder production by spray bed drying

16.1 The Spray Drying Process

Spray drying converts a liquid feed — solution, suspension, or emulsion — into a dry powder in a single, continuous step. The liquid is atomized into fine droplets (50–200 μm) by a rotary atomizer or two-fluid nozzle, and the droplets are introduced into a hot air chamber (inlet temperature 150–250°C) where rapid evaporation (< 5 seconds residence time) dries each droplet to a particle before it reaches the chamber wall. Product temperature typically reaches only 40–80°C despite the high inlet air temperature, because the evaporative cooling of the droplet limits the surface temperature during drying.

Spray drying produces powders with controlled properties: moisture content (typically 1–5%), particle size (10–200 μm , controlled by atomizer speed or nozzle pressure), particle morphology (solid, hollow, or porous spheres), and — for encapsulated products — core material retention and encapsulation efficiency.

16.2 Microencapsulation for Sensitive Ingredients

Spray drying microencapsulation is the primary method for protecting nutraceutical ingredients against oxidation (omega oils, fat-soluble vitamins), moisture (hygroscopic botanicals), heat (probiotics in low-inlet-temperature spray drying), light (carotenoids, chlorophyll), and GI degradation (probiotics, enzymes). The ingredient (core material) is emulsified or dispersed in an aqueous solution of the wall material (typically modified starch, maltodextrin, gum arabic, whey protein, or their combinations), and the emulsion is spray dried.

Core Material	Wall Material System	Critical Parameter	Target Product Spec
Fish oil (omega-3)	Modified starch + WPC	Emulsification droplet size < 1 μm	Peroxide value < 5 meq/kg; encap. efficiency > 90%
Probiotic cultures	Skim milk / maltodextrin	Inlet temp $\leq$ 140°C; outlet $\leq$ 70°C	Survival $\geq$ 80%; target CFU/g $\pm$ 0.3 log

Spray-dried botanical extract	Maltodextrin 10–20 DE	Feed total solids 20–40%	Moisture $\leq$ 5%; marker compound retention $\geq$ 95%
Vitamin D3 (oil solution)	Starch + tocopherols (antioxidant)	O2 exclusion in feed preparation	Potency $\geq$ 95% of label; peroxide value < 3
Curcumin (lipophilic)	HPMC + lecithin	Emulsion stability before spray drying	Water dispersibility OD > 0.8 at 1 mg/mL

PROC-X EQUIPMENT SPOTLIGHT Proc-X spray drying systems for nutraceutical applications range from 5 kg/hr evaporative capacity (laboratory/development) to 500 kg/hr (production). All units include an oxygen-depleted drying chamber option (N2 or CO2 inlet gas) for oxidation-sensitive materials, integrated cyclone and bag filter, CIP-capable atomizer, and PX-BMS data logging. The PX-SDB-Agglomerate models add a fluid bed bottom section for simultaneous agglomeration, producing instant-dispersing granules directly from the spray dryer — eliminating a separate granulation step.

Chapter 17: Botanical Extract Processing and Standardization

IN THIS CHAPTER

- ▶ The botanical supply chain challenge: natural variability and its consequences
- ▶ Extraction processes: percolation, maceration, supercritical CO₂
- ▶ Spray drying and vacuum belt drying for extract concentration
- ▶ Marker compound standardization: milling and blending to specification
- ▶ Quality systems for botanical identity: authentication vs. adulteration

17.1 The Natural Variability Challenge

Botanical raw materials present a processing challenge unique to the nutraceutical industry: extreme lot-to-lot variability in the concentration of the active marker compound, driven by differences in growing region, harvest season, plant age, agricultural practices, and post-harvest handling. A standardized elderberry extract specified at 13% anthocyanins may arrive from different suppliers or lots ranging from 8% to 25% anthocyanins — all meeting generic 'elderberry extract' identity testing, but requiring very different blend ratios to achieve the label claim.

The processing solution is standardization blending: characterize each lot for marker compound content, then blend lots (or blend a high-marker lot with inert diluent) to achieve the target concentration with a defined specification window. Proc-X V-blenders and bin blenders with inline NIR monitoring enable the statistical validation of blend uniformity needed to support standardization claims.

17.2 Milling Botanical Extracts

Spray-dried botanical extracts are typically produced as fine powders ($D_{90} < 200 \mu\text{m}$) from the spray dryer. However, some botanical materials — dried whole plant material, root powders, bark powders — require milling before extraction or before encapsulation. Proc-X hammer mills (for fibrous botanical materials resistant to impact milling) and oscillating granulators (for friable spray-dried extracts) are the primary botanical milling tools.

- Fibrous materials (valerian root, ashwagandha root, turmeric): Hammer mill at 3,000–4,500 rpm through 0.5–1.5 mm screen. Target $D_{90} < 500 \mu\text{m}$ for extraction efficiency or capsule fill uniformity.
- Spray-dried extracts: Oscillating granulator or Comil at low speed to break up cakes without heat generation or electrostatic charging of the fine powder.

REGULATORY CONTEXT *Authentication of botanical raw materials is a regulatory requirement under 21 CFR Part 111.75 — every dietary supplement manufacturer must verify the identity of each component before use. The FDA expects this to include a scientifically valid identity test, not merely visual inspection. For standardized extracts, identity testing typically uses HPLC for marker compound quantitation, combined with*

FTIR or DNA barcoding for species verification. Processing cannot compensate for a misidentified raw material — robust incoming quality control is the first line of defense.

Chapter 18: Encapsulation, Softgel & Delivery System Processing

IN THIS CHAPTER

- ▶ Hard capsule filling: types, fill weights, and cGMP considerations
- ▶ Softgel encapsulation: gelatin vs. non-gelatin shells and oil-fill processing
- ▶ Modified release bead and pellet production for capsule fill
- ▶ Liquid-filled hard capsule technology for BCS Class II APIs

18.1 Hard Capsule Filling

Hard capsule filling involves dosing a measured quantity of blend, granule, pellet, or powder into a two-piece hard-shell capsule (gelatin or HPMC) and closing the capsule. The capsule filling machine — either a dosing disc or tamping pin machine for powders, or a vibrating tray for pellets/tablets — is the final size-controlling step in the solid dosage manufacturing chain. Critical quality attributes are weight variation ($\pm 5\%$ of target for most products, $\pm 3\%$ for potent APIs) and content uniformity.

Upstream processing directly determines capsule filler performance: if the blend has poor flow, the capsule filler will have high weight variability. Proc-X processing systems optimize the upstream steps (granulation, blending, milling) specifically for the capsule filling step, using the flow function coefficient (FFC) and basic flowability energy (BFE) as the measured endpoints of upstream processing to ensure fill weight consistency.

18.2 Softgel Encapsulation

Softgel capsules consist of a one-piece gelatin or non-gelatin (HPMC-based Vcaps soft) shell enclosing a liquid, suspension, or paste fill. The primary application in nutraceuticals is oil-based materials: omega-3 fish oils, vitamin D3 in oil, vitamin E, fat-soluble botanical extracts, and CBD oil. The Rotary Die Encapsulation process forms and fills the capsule simultaneously — gelatin ribbon is formed on two counter-rotating die rolls, fill is injected between the ribbons as the dies close, and filled sealed capsules are cut from the ribbon.

Critical processing steps upstream of the encapsulation machine: API or ingredient dispersion in the fill oil (using a high-shear mixer or rotor-stator), nitrogen sparging to reduce dissolved oxygen (preventing oxidation during encapsulation), and de-aeration of the fill mass to prevent bubbles in the sealed capsule.

PROC-X EQUIPMENT SPOTLIGHT Proc-X high-shear dispersers for softgel fill preparation are equipped with vacuum de-aeration as standard — the fill vessel is evacuated to < 50 mmHg during mixing to remove dissolved air before the fill mass is transferred to the encapsulation machine. Dissolved oxygen in the fill oil is one of the primary drivers of omega-3 oxidation during shelf life. Proc-X systems achieve dissolved oxygen levels < 1 ppm in the fill mass before encapsulation.

PART V

STERILE & INJECTABLE

Aseptic Processing, Lyophilization & Sterile API Milling

Chapter 19: Aseptic Processing Principles and Equipment

IN THIS CHAPTER

- ▶ The sterility assurance level (SAL) and why 10^{-6} is the pharmaceutical standard
- ▶ Aseptic mixing equipment: closed, steam-sterilizable vessels
- ▶ Environmental monitoring and cleanroom classifications (ISO 5 through 8)
- ▶ Single-use vs. stainless steel for sterile mixing: the trade-offs
- ▶ Mixing validation for sterile drug products: mixing uniformity and time studies

19.1 Why Sterile Manufacturing Requires a Different Approach

Sterile drug products — injectables, ophthalmic solutions, inhalation solutions, and lyophilized powders — are administered by routes (IV, IM, SQ, intrathecal) that bypass the body's microbial defense systems. A non-sterile parenteral product can cause sepsis, meningitis, or death. The regulatory and engineering consequences of this risk are profound: every equipment surface that contacts the sterile drug product must be sterilizable, every connection must be aseptic, and every manufacturing operation must be performed in a classified cleanroom with validated environmental monitoring.

The FDA's Aseptic Processing Guidance (2004) and the EU GMP Annex 1 (2022 revision) establish the current regulatory expectations for aseptic manufacturing. The 2022 Annex 1 revision placed Contamination Control Strategy (CCS) at the center of the regulatory framework — manufacturers must demonstrate that every potential contamination source has been identified, assessed, and controlled. Processing equipment is the most significant contamination risk in aseptic manufacturing, and equipment design is the primary mitigation.

19.2 Sterile Mixing Equipment

Mixing operations for sterile drug products include: dissolution of API in sterile water for injection (WFI), blending of lyophilized formulation components, compounding of concentrated solutions before sterile filtration, and pH adjustment. Equipment requirements:

- **Sterilization:** All product-contact surfaces must be sterilizable. Methods: steam sterilization (SIP — sterilize-in-place, $121^{\circ}\text{C} \times 30 \text{ min}$), or H_2O_2 vapor sterilization for single-use components. Proc-X sterile mixing vessels have full SIP capability with validated drain points and steam distribution.
- **Enclosed design:** No open-top mixing vessels in ISO 5 cleanrooms. All product contact via closed, sealed vessels with aseptic connections (tri-clamp, aseptic

coupling). No gaskets or O-rings in product zones that cannot be validated for extractables/leachables.

- Surface finish: $Ra \leq 0.5 \mu\text{m}$ in product zones (vs. $Ra \leq 0.8 \mu\text{m}$ for oral solid). Electropolished 316L SS. Passivated per ASTM A967. Full boroscope weld inspection with documentation.
- Mixing mechanism: Magnetic coupling (preferred for sterile vessels — eliminates shaft seals that are contamination risk points) or triple mechanical seal for larger vessels.

PROC-X EQUIPMENT SPOTLIGHT Proc-X sterile mixing vessels (PX-SV series) use bottom-mounted magnetic drive mixers as standard for vessels up to 2,000 L — eliminating the top mechanical seal that is the single most common source of particulate contamination in traditional sterile mixing vessels. The magnetic drive has zero dynamic seals in the product zone. Magnets are encapsulated in 316L SS. Full SIP capability validated to $F0 \geq 12$ at all product-contact surfaces.

19.3 Single-Use vs. Stainless Steel for Sterile Mixing

Criterion	Single-Use (SUB)	Stainless Steel (Fixed)	Recommendation
Capital cost	Lower (no vessel purchase)	Higher	SUS for clinical; SS for commercial
Cleaning validation	Not required (disposable)	Required (full validation package)	SUS advantage for multi-product
Sterilization	Pre-sterilized (gamma or EO)	SIP required each use	SUS simpler for batch-to-batch
Extractables/leachables	Higher risk (polymer contact)	Lower risk (316L SS, established profiles)	SS advantage for high-risk formulations
Scale range	50–2,000 L (most commercial)	1–50,000 L	SS advantage for large volume
Turnaround time	Faster (no cleaning/SIP cycle)	Slower (cleaning + SIP + release)	SUS advantage for fast-turn clinical

Chapter 20: Sterile API Milling and Micronization

IN THIS CHAPTER

- The added complexity of milling an API that must remain sterile
- Sterile jet milling: equipment, environmental control, and validation
- Micronization of inhalation APIs: the DPI particle size imperative
- Contained sterile milling for high-potency cytotoxics

20.1 The Sterile Milling Challenge

Milling an API that will be used in a sterile drug product without a terminal sterilization step requires the milling operation to be performed aseptically — the API must be sterile before milling, the milling equipment must be sterilized, and the entire milling operation must be conducted in a classified environment (ISO 7 or better) with validated environmental monitoring. This is rare — most APIs are terminally sterilized by gamma irradiation after milling — but for gamma-sensitive APIs (proteins, peptides, oligonucleotides), aseptic milling is required.

20.2 DPI API Micronization

Dry powder inhalers (DPIs) require API particles with MMAD 1–5 μm for adequate lung deposition. The particle size specification is among the tightest in all of pharmaceutical manufacturing — a shift in MMAD from 2.5 μm to 3.5 μm can reduce fine particle fraction (FPF, the fraction of the dose that deposits in the lung) by 30–40%, dramatically reducing clinical efficacy. Jet milling parameters for DPI API micronization are defined tightly and controlled with inline monitoring:

DPI API	MMAD Target	Classifier Setting	Critical Control
Budesonide (corticosteroid)	1.5–2.5 μm	12,000–14,000 rpm	N ₂ feed; humidity < 10% RH in mill zone
Salbutamol sulfate (bronchodilator)	2.0–3.5 μm	9,000–12,000 rpm	Crystal form verification before and after milling
Tiotropium bromide (LAMA)	1.8–2.8 μm	11,000–13,000 rpm	OEB 4 containment; N ₂ inerting
Fluticasone propionate	1.5–2.5 μm	12,000–15,000 rpm	Electrostatic charge control; humidity monitoring

ENGINEERING INSIGHT *Electrostatic charge is the single greatest source of variability in DPI API jet milling. Fine particles (< 5 μm) generated by jet milling accumulate electrostatic charge that causes them to adhere to equipment surfaces and to the lactose carrier particles in the DPI formulation — changing the aerodynamic performance of the finished product. Proc-X DPI jet mill systems include static-dissipative linings on all product contact surfaces, nitrogen (< 1% RH) carrier gas, and anti-static grounding of the cyclone and filter systems.*

Chapter 21: Lyophilization Preparation — Formulation Mixing and Filling

IN THIS CHAPTER

- ▶ Lyophilization (freeze-drying) overview: why it's chosen and what it demands of upstream processing
- ▶ Formulation compounding for lyophilized injectables
- ▶ Fill accuracy, stopper placement, and the role of mixing in lyo success
- ▶ PAT for lyo primary drying endpoints

21.1 Why Lyophilize?

Lyophilization (freeze-drying) converts a liquid drug formulation into a stable dry cake by freezing the solution and then subliming the ice under vacuum. It is chosen when: (1) the drug is unstable in aqueous solution with insufficient shelf life; (2) the drug degrades at elevated temperatures, excluding spray drying; (3) the drug has a very short reconstitution window and must be delivered as a powder. Biopharmaceuticals (monoclonal antibodies, enzymes, peptide hormones) and chemically labile small molecules (vancomycin, cephalosporins, proton pump inhibitors) are the primary lyo candidates.

The formulation mixing step before fill-finish is critical to lyophilization success. The excipients — bulking agents (mannitol, sucrose, trehalose), lyoprotectants (sucrose, trehalose), cryoprotectants (glycerol, PEG), and buffering agents — must be dissolved completely and uniformly distributed before filling. Any concentration gradient in the filled vials will produce non-uniform cake structure and potentially non-uniform API content. Proc-X sterile mixing vessels with integrated conductivity and pH monitoring ensure complete dissolution and homogeneity before each fill.

21.2 Fill Accuracy Requirements

For lyophilized injectables, fill weight accuracy is directly linked to the dose accuracy of the reconstituted product. USP <1> requires that the volume of injections meet the labeled volume when reconstituted. Overfill requirements (e.g., 10 mg/vial with 10% overfill for reconstitution losses) mean the fill weight must meet a tight window. For high-potency lyophilized oncology drugs, fill weight specification may be $\pm 1\text{--}2\%$ of target. Upstream mixing homogeneity determines whether this specification is achievable.

PROC-X EQUIPMENT SPOTLIGHT Proc-X sterile compounding vessels for lyo preparation use a mixing verification protocol (PX-MVP) that monitors conductivity and API concentration via inline UV or NIR throughout the compounding step, with automatic endpoint detection when conductivity and concentration have stabilized within specification. The PX-MVP eliminates the offline sampling-and-analysis cycle that has historically added 2–4 hours to the lyophilization preparation timeline.

PART VI

EQUIPMENT DEEP DIVES

The Machines Behind the Medicine

Chapter 22: High-Shear Granulators and Fluid Bed Processors

IN THIS CHAPTER

- ▶ Complete design anatomy of the high-shear granulator — every component explained
- ▶ The fluid bed processor control architecture and how it manages the four critical process variables
- ▶ Scale-up from 2 L to 300 L: the data you need and the equipment choices that matter
- ▶ Combination systems: one-pot processors integrating granulation, drying, and coating

22.1 High-Shear Granulator Design Anatomy

A complete high-shear granulator consists of the following subsystems, each critical to cGMP operation:

- Bowl and lid assembly: 316L SS, electropolished, jacketed. Lid seals with inflatable EPDM gasket that can be sterilized and integrity-tested. All lid penetrations (impeller shaft, chopper shaft, spray nozzle, sampling port, exhaust) are sealed with FDA-compliant elastomers.
- Impeller drive: Top-mounted or bottom-mounted (bottom-mounted preferred for contained discharge). AC induction motor with VFD providing 0.5–100 rpm in low speed (discharge) and 100–600 rpm in granulation mode. Torque sensor on drive shaft, resolution 0.1% of full scale.
- Chopper drive: Side-mounted, 0.5–4 kW motor, 1,000–3,600 rpm via VFD. Chopper blade geometry: 4-blade flat bar (standard) or 3D profiled blade for sensitive granules.
- Spray system: Peristaltic pump (0.1–2,000 mL/min range) driving a binary atomizing nozzle (compressed air + liquid streams). Droplet size $D(4,3)$ 50–150 μm at standard operating conditions. Proc-X systems include gravimetric spray verification — the binder vessel is weighed in real time and the PLC compares actual liquid added against the recipe setpoint.
- CIP spray balls: Two 360° rotating spray balls mounted on the lid, providing full coverage of all internal surfaces with cleaning solution at ≥ 5 psi impact pressure.

22.2 Fluid Bed Processor Control Architecture

The fluid bed processor control system must manage four interdependent variables simultaneously: product temperature, moisture content, fluidization quality (bed pressure drop and visual bed stability), and spray rate. Proc-X PX-FBP control architecture:

- Cascade control: Spray rate is the manipulated variable in a cascade control loop where the primary controller is product temperature (or outlet air dew point for moisture-endpoint control). When product temperature falls below setpoint

(insufficient drying), spray rate decreases automatically. This prevents overwetting without operator intervention.

- Fluidization control: Differential pressure across the product bed is maintained in the target range (10–50 mm H₂O) by adjusting the supply air fan speed via VFD. Low differential pressure (loss of fluidization) triggers an alarm and spray rate cutoff before the bed collapses.
- Process endpoint control: LOD (loss on drying) endpoint is determined either by offline sampling (traditional) or by inline NIR measurement of moisture in the product bed (PAT approach). NIR endpoint reduces drying time variation by $\pm 30\%$ versus time-endpoint control.

PROC-X EQUIPMENT SPOTLIGHT The Proc-X PX-ONE one-pot processor combines high-shear granulation and fluid bed drying in a single vessel. After high-shear wet granulation, the impeller is retracted and process air is introduced through the perforated base plate to fluidize and dry the wet granules without transfer. Eliminates the wet granule transfer step that is a primary source of granule attrition, segregation, and contamination risk. Available in 10, 25, 65, and 150 L sizes.

Chapter 23: Pharmaceutical Mills — From Comil to Jet Mill

IN THIS CHAPTER

- ▶ The complete pharmaceutical mill taxonomy: when to use each type
- ▶ Conical mills (Comils): the workhorse for granule milling and deagglomeration
- ▶ Pin mills and hammer mills: high-throughput dry size reduction
- ▶ Oscillating granulators: gentle milling for fragile granules
- ▶ The complete mill selection decision tree

23.1 The Pharmaceutical Mill Taxonomy

Mill Type	Mechanism	Product D90	Throughput	Key Pharma Application
Conical mill (Comil)	Impeller + screen	0.5–3 mm	5–2,000 kg/hr	Granule milling after HSG/RC, deagglomeration
Oscillating granulator	Oscillating bar + screen	1–5 mm	20–500 kg/hr	Gentle milling of friable granules, slugs
Pin mill	High-speed pin disc impact	50–500 μ m	50–5,000 kg/hr	Spray-dried powders, botanical extracts
Hammer mill	High-speed hammer impact	100 μ m – 5 mm	100–10,000 kg/hr	Botanical powders, API pre-milling
Air classifier mill (ACM)	Impact + integrated classif.	10–200 μ m	10–500 kg/hr	Fine API milling, excipient particle sizing
Jet mill (fluidized bed)	Particle-particle impact	1–50 μ m	0.5–500 kg/hr	API micronization, DPI formulations
Bead mill (wet)	Media-particle attrition	0.1–10 μ m	1–1,000 L/hr	API nanosizing, wet grinding

23.2 Conical Mills — the Pharma Workhorse

The conical mill (Comil) is the single most widely used size reduction device in solid dose pharmaceutical manufacturing. Its rotating impeller forces material through a conical screen, and the gap between impeller and screen controls the product particle size. Variable parameters: impeller type (round, square, grater-style), impeller speed (200–3,600 rpm), and screen aperture (0.075–6.0 mm). Advantages: gentle milling action (preserves granule morphology), easy screen changeover, and contained design for OEB 3–4 applications. Proc-X Comil equivalents offer continuous product temperature monitoring and screen integrity monitoring via differential pressure.

23.3 Mill Selection Decision Logic

Selecting the right pharmaceutical mill for a given application depends on: (1) target particle size, (2) API sensitivity (thermal, mechanical, electrostatic), (3) throughput

requirement, (4) batch vs. continuous operation, (5) containment requirement (OEB classification), and (6) downstream process requirements (flowability, bulk density). When target D90 is $> 500 \mu\text{m}$ and the material is not fragile, a hammer mill or Comil is appropriate. When target D90 is $50\text{--}500 \mu\text{m}$ and heat is a concern, an air classifier mill with cooling option is preferred. When target D90 is $< 50 \mu\text{m}$, a jet mill with nitrogen gas is required. When the suspension is the desired form, a wet bead mill achieves the finest particle sizes.

Chapter 24: Twin-Screw Extruders and Continuous Processing Equipment

IN THIS CHAPTER

- ▶ TSE design fundamentals for pharmaceutical HME and continuous wet granulation
- ▶ Scale-up from 11 mm to 40 mm: the geometric similarity principles
- ▶ Continuous wet granulation by TSE: equipment, process, and PAT integration
- ▶ Downstream HME processing: cooling conveyors, pelletizers, and milling

24.1 TSE Design for Pharmaceutical Applications

The pharmaceutical co-rotating twin-screw extruder differs from its industrial counterpart in several critical ways: all product-contact surfaces are 316L SS or GMP-grade polymer (PEEK, PTFE); the barrel is modular with individual temperature-controlled zones; a vacuum vent is available to remove trapped air or residual solvent from the melt; and the screw is fully removable for cleaning and inspection without dismantling the entire gearbox assembly.

Screw designs for pharmaceutical HME are custom-developed using process simulation software (PolyFlow, SIGMA, TSIM) that predicts melt temperature, shear rate, and residence time distribution for each screw configuration. Proc-X supplies reference screw configurations for the most common HME applications (HPMC-AS ASD, PVP-VA ASD, enteric matrix extrusion) and custom screw design services for novel formulations.

24.2 Continuous Wet Granulation by TSE

In continuous wet granulation (CWG), powder and granulating liquid are continuously fed into a co-rotating TSE. Granulation occurs in the wetted zone of the screw barrel in 20–60 seconds. Key advantages over batch high-shear granulation:

- No scale-up: The same screw configuration operates from development (< 1 kg/hr at 11 mm) to production (> 100 kg/hr at 40–50 mm). The TSE operates continuously at all scales — scale-up is achieved by increasing throughput rate, not by changing process geometry.
- Narrow residence time distribution: Batch granulators have broad within-batch variability in granule processing time. TSE RTD is typically < ± 5 seconds, producing much more uniform granule properties.
- Real-time process control: Inline PAT instruments (FBRM, NIR) provide continuous feedback, enabling closed-loop control of liquid-to-solid ratio and granule size in real time.

PROC-X EQUIPMENT SPOTLIGHT Proc-X PX-CWG continuous wet granulation lines integrate the PX-TSE granulator with: dual gravimetric feeders (powder + liquid), inline FBRM probe for granule size monitoring, a purpose-designed transfer chute to the PX-

FBP fluid bed dryer, and the PX-BMS data system — all controlled from a single HMI. The complete line occupies 8 m × 3 m floor space and operates from 5 kg/hr (development) to 150 kg/hr (full commercial production) on the same equipment frame.

Chapter 25: Bulk Material Handling in Pharmaceutical Facilities

IN THIS CHAPTER

- ▶ Contained powder transfer for high-potency APIs — OEB 1 through OEB 5
- ▶ Vacuum conveying for pharmaceutical facilities: design and validation
- ▶ Bin (IBC) management systems in continuous and batch manufacturing
- ▶ Loss-in-weight feeders: the critical link in continuous manufacturing

25.1 Contained Powder Transfer for High-Potency APIs

High-potency APIs (HPAPIs) — cytotoxics, hormones, cardiovascular agents, and oncology compounds — require contained handling to protect operators from occupational exposure. The containment approach scales with the OEB classification:

OEB Level	OEL Range	Transfer Approach	Proc-X Solution
OEB 1	> 1 mg/m ³	Standard engineering controls, local exhaust ventilation	Standard IBC charging, ventilated booth
OEB 2	0.1–1 mg/m ³	Closed charging, dust extraction	Split butterfly valve, closed charging station
OEB 3	10–100 µg/m ³	Contained charging isolator or glove bag	PX-CBV contained charging with HEPA exhaust
OEB 4	1–10 µg/m ³	Fully enclosed isolator, HEPA filtered	PX-ISO containment isolator (< 1 µg/m ³ verified)
OEB 5	< 1 µg/m ³	Highest containment — dedicated facility/equipment	PX-HPISO (< 0.1 µg/m ³ with continuous air monitoring)

25.2 Vacuum Conveying in Pharmaceutical Facilities

Vacuum (dilute-phase pneumatic) conveying is the preferred method for transferring pharmaceutical powders between processing steps within a facility. Advantages over mechanical conveying: completely enclosed (no open powder exposure), gentle transport (suitable for granules that would attrit in mechanical conveyors), and easy cleaning (piping can be CIP or WIP cleaned). Proc-X pharmaceutical vacuum conveying systems use: food-grade polyurethane or PTFE-lined stainless steel tubing, gravity-assisted vertical drops where possible (minimizes particle attrition), and a pulsed compressed air cleaning cycle that dislodges product from pipe walls between batches.

25.3 Loss-in-Weight Feeders for Continuous Manufacturing

Loss-in-weight (LIW) feeders are the heart of any continuous pharmaceutical manufacturing line. They provide the precise, continuous mass flow rates needed to

maintain constant drug-to-excipient ratios in a continuous process. Performance requirements:

- Feed accuracy: $\pm 0.5\%$ of setpoint for all materials in the formulation (API, excipients, binders).
- Refill strategy: Automated refill from overhead IBC without significant flow rate disturbance. Proc-X PX-LIW feeders use fast-refill (< 30 second refill cycle) with feedforward compensation that maintains flow rate accuracy during and after refill.
- API segregation monitoring: Low-dose APIs (< 5% of formulation) require independent verification that the API feeder is actually delivering. Inline Raman or NIR spectroscopy confirms API presence in the blend stream after mixing.

ENGINEERING INSIGHT *The refill transition in a continuous process is the single highest-risk event for product quality. During the 15–60 seconds required to refill a LIW feeder hopper, the feeder briefly transitions from gravimetric (closed-loop weight-based) to volumetric (open-loop speed-based) mode. If the bulk density of the material in the refill batch differs from the previous batch — which occurs routinely with natural variation in excipient lots — the volumetric mode will deliver an incorrect flow rate during the transition. Proc-X LIW feeders use a proprietary fast-refill algorithm that characterizes the bulk density of each refill batch automatically and updates the volumetric mode speed setting before the transition occurs.*

PART VII

COMPLIANCE, QUALITY & THE DIGITAL FUTURE

Validation, Data Integrity & Continuous Manufacturing

Chapter 26: 21 CFR Part 211 — Complete cGMP Compliance for Processing Equipment

IN THIS CHAPTER

- ▶ Every equipment-relevant section of 21 CFR Part 211 — requirements and Proc-X compliance approach
- ▶ The FDA's 483 observation and warning letter landscape for equipment
- ▶ How the Quality by Design (QbD) framework changes equipment specification
- ▶ Pharmaceutical Inspection Co-operation Scheme (PIC/S) and ICH Q10 alignment
- ▶ Building a compliance file: the documentation Proc-X provides with every system

26.1 Complete 21 CFR Part 211 Equipment Requirements

21 CFR Part 211 Subpart D (Sections 211.63 through 211.68 and 211.182) governs equipment in pharmaceutical manufacturing facilities. This chapter provides a section-by-section analysis of every equipment-relevant requirement and the specific Proc-X design response to each.

Section 211.63 — Equipment Design, Size, and Location

Requirement: Equipment must be of appropriate design, adequate size, and suitably located to facilitate operations for its intended use and for its cleaning and maintenance.

Proc-X approach: Every Proc-X system is delivered with a Design Qualification (DQ) document that records the equipment specification rationale — why the bowl size was selected for the batch range, why the impeller geometry was chosen for the material properties, and why the equipment location in the facility layout supports cleaning and maintenance access. The DQ is part of the Equipment History File (EHF) maintained throughout the equipment lifecycle.

Section 211.65 — Equipment Construction

Requirement: Surfaces that contact components, in-process materials, or drug products must not be reactive, additive, or absorptive so as to alter the safety, identity, strength, quality, or purity of the drug product.

Proc-X approach: All product-contact materials are documented in the Material of Construction (MoC) Certificate provided with every system. Primary materials: ASTM A270 316L SS for metal surfaces; FDA 21 CFR 177.2600-compliant elastomers (EPDM, silicone, PTFE) for seals and gaskets; borosilicate glass for sight glasses; PEEK or PTFE for bearings in product zones where metal contact is unacceptable. No cadmium, lead, or hexavalent chromium plating on any surface. No painted surfaces in product zones.

Section 211.67 — Equipment Cleaning and Maintenance

Requirement: Equipment shall be cleaned, maintained, and sanitized at appropriate intervals to prevent malfunctions or contamination that would alter the safety, identity, strength, quality, or purity of the drug product.

Proc-X approach: (1) Cleanability by design — all internal radii ≥ 3 mm; no dead legs in piping; spray ball coverage verified by riboflavin fluorescent dye test documented in the IQ package. (2) CIP/WIP design — all Proc-X systems include CIP circuits as standard. (3) Cleaning SOP — Proc-X provides a draft cleaning SOP based on equipment geometry for customer validation adaptation. (4) Maintenance access — all bearings, seals, and wear parts accessible without product zone disassembly; PM schedule in the IQ documentation.

Section 211.68 — Automatic, Mechanical, and Electronic Equipment

Requirement: Computerized systems used to perform GMP functions must be validated, calibrated, and protected against unauthorized changes. Input and output verification must be performed at a frequency appropriate to the complexity and reliability of the system.

Proc-X approach: All PX-BMS control systems are validated under GAMP 5 Category 4 (configured product) protocols. System includes: role-based access control with password authentication, complete electronic audit trail (who changed what, when, and why), electronic batch records locked after batch completion, and calibration management module that tracks all instrument calibration schedules and alerts to upcoming calibrations.

Section 211.182 — Equipment Cleaning and Use Log

Requirement: Written records shall be maintained for equipment used in the manufacture, processing, packing, or holding of a drug product. Records shall include: date, time, product, lot number, operator name for each use; cleaning and sanitizing performed; and identification of the person who performed these activities.

Proc-X approach: PX-BMS automatically generates the equipment use log for every batch — date, time, operator ID (from electronic login), product, batch number, process parameters, and cleaning confirmation (with electronic signature from cleaning operator). The log is part of the electronic batch record and is retained indefinitely in the historian.

REGULATORY INSIGHT *The FDA issued 68 warning letters citing equipment-related GMP violations in the most recent three-year reporting period — primarily related to data integrity failures in computerized equipment systems (manipulated audit trails, shared logins, backdated records) and inadequate cleaning validation. Proc-X GAMP 5-validated control systems are designed specifically to make these violations impossible: audit trails cannot be disabled, logins cannot be shared, and timestamps are generated by a synchronized network clock that cannot be altered by any user at any access level.*

26.2 The QbD Framework and Equipment Specification

Quality by Design (ICH Q8) connects product quality to process parameters through a structured development program. For processing equipment, QbD changes the equipment specification approach from 'specify a mixer and validate it works' to 'define the design space for mixing, specify equipment capable of operating throughout the design space

with measurement of all critical parameters, and demonstrate the equipment's control capability across the design space.'

Proc-X equipment is specified to be 'design space aware': every controllable process parameter (speed, temperature, time, vacuum, spray rate) has a validated operating range documented in the equipment's OQ protocol, and the ranges are cross-referenced to the formulation design space established in the QbD development program. When the process design space changes during development, the equipment OQ update is straightforward — because the relationship between equipment parameters and product quality was built in from the start.

26.3 The Compliance Documentation Package

Document	Content	Regulatory Basis
Design Qualification (DQ)	Equipment design rationale vs. URS requirements	21 CFR 211.63, GAMP 5
Factory Acceptance Test (FAT)	Function testing at Proc-X facility pre-shipment	ICH Q10, PIC/S PI 006
Site Acceptance Test (SAT)	Function testing at customer site post-installation	ICH Q10, PIC/S PI 006
IQ Protocol & Report	Installation correctness — utilities, materials, documentation	21 CFR 211.68, GAMP 5
OQ Protocol & Report	Equipment operates per spec across full operating range	21 CFR 211.63, GAMP 5
PQ Protocol & Report	Process + equipment consistently produces conforming product	21 CFR 211.100, ICH Q10
Material of Construction (MoC) Certificate	All product-contact materials documented	21 CFR 211.65
Surface Finish Certificate	Profilometer measurement of all product-contact surfaces	21 CFR 211.65
Calibration Records	Pre-delivery calibration of all instruments	21 CFR 211.68
Weld Certificates	Boroscope inspection and passivation records for all welds	21 CFR 211.65
Cleaning Validation Support Package	CIP spray ball coverage study, riboflavin test results	21 CFR 211.67
Computer System Validation (CSV) Package	GAMP 5 Category 4 validation documentation for PLC/SCADA	21 CFR 11, EU Annex 11

Chapter 27: Process Validation — Stage 1, Stage 2, and Stage 3

IN THIS CHAPTER

- ▶ FDA Process Validation Guidance (2011): the three-stage lifecycle model
- ▶ Stage 1 — Process Design: where equipment performance data feeds into development
- ▶ Stage 2 — Process Qualification: PPQ batches and statistical acceptance criteria
- ▶ Stage 3 — Continued Process Verification: the SPC and CPV program
- ▶ Equipment performance verification within the continued process verification framework

27.1 The Three-Stage Validation Lifecycle

The FDA's 2011 Process Validation Guidance replaced the traditional three-batch validation concept with a continuous, lifecycle-based approach. The guidance recognizes that a process cannot be considered validated after three batches at a fixed set of parameters — rather, validation is an ongoing activity that encompasses process development, qualification, and continued monitoring throughout the commercial lifecycle.

Stage 1 — Process Design

Process design uses the knowledge gained during pharmaceutical development (QbD, design of experiments, risk assessment) to establish the commercial manufacturing process. For processing equipment, Stage 1 activities include: equipment characterization (mapping of temperature uniformity in dryers, blending kinetics in blenders, milling PSD as a function of classifier speed), design space definition (the range of equipment parameters within which the process produces conforming product), and failure mode analysis (PFMEA identifying potential equipment-related failure modes before they occur in commercial manufacturing).

Stage 2 — Process Performance Qualification (PPQ)

PPQ confirms that the commercial manufacturing process, as defined in Stage 1, performs as expected and produces a product meeting all CQAs. For processing equipment, PPQ requires demonstrating equipment performance under commercial-scale conditions using commercial-grade raw materials, commercial batch sizes, and the commercial process parameters defined in the process description. Statistical acceptance criteria (typically $n \geq 3$ consecutive conforming batches, with $Cpk \geq 1.33$ for all CQAs) confirm that the process is in control.

Stage 3 — Continued Process Verification (CPV)

CPV is the ongoing monitoring program that provides statistical evidence that the process remains in a state of control throughout commercial manufacturing. For equipment, CPV monitors the relationship between critical process parameters (CPPs) and critical quality attributes (CQAs) using statistical process control. Proc-X PX-BMS provides the CPV data

infrastructure: all batch data is automatically exported to the customer's statistical analysis system, and Shewhart control charts (X-bar/R charts for CPPs) and capability indices (Cpk for CQAs) are updated after each batch.

PROC-X EQUIPMENT SPOTLIGHT Proc-X CPV support package includes 24 months of trending analysis on all PX-BMS-logged process parameters, generating monthly CPV reports that identify parameter drifts before they cause product failures. For facilities without a dedicated statistical analyst, Proc-X offers managed CPV services: monthly reports prepared by Proc-X validation engineers, reviewed with the customer's quality team, and filed in the customer's quality management system. This service has prevented 7 product quality deviations by detecting equipment parameter drift in the early stages across our customer base.

Chapter 28: Data Integrity, 21 CFR Part 11, and EU Annex 11

IN THIS CHAPTER

- ▶ The ALCOA+ framework: Attributable, Legible, Contemporaneous, Original, Accurate — plus Complete, Consistent, Enduring, Available
- ▶ 21 CFR Part 11 requirements for electronic records and electronic signatures
- ▶ EU GMP Annex 11 — computerised systems in pharmaceutical manufacturing
- ▶ Common data integrity failures in processing equipment and how to prevent them
- ▶ GAMP 5 Category 4 validation for PLC-based pharmaceutical equipment

28.1 The Data Integrity Imperative

Data integrity — the completeness, consistency, and accuracy of GMP data throughout its lifecycle — has become the FDA's and EMA's top inspection focus. Between 2014 and 2024, data integrity failures were cited in the majority of FDA import alerts and warning letters against pharmaceutical manufacturers, particularly those in Asia. The consequences are severe: import alerts, consent decrees, facility shutdowns, and criminal prosecution of company executives.

For processing equipment, data integrity requirements mean: every instrument reading must be recorded by the system itself (not manually transferred); audit trails must capture all changes to electronic records; electronic signatures must meet 21 CFR Part 11 requirements; and no single user may have the ability to both create and approve a record (segregation of duties).

28.2 ALCOA+ in Processing Equipment Systems

- **Attributable:** Every data record must identify the person or system that created it. Proc-X PX-BMS requires individual user login for all data entry; shared generic logins are architecturally impossible.
- **Legible:** Records must be readable throughout the retention period. PX-BMS stores records in an open-standard format (CSV + XML) readable by standard tools, not proprietary binary formats that become inaccessible when software versions change.
- **Contemporaneous:** Records must be created at the time of the activity. PX-BMS timestamps all records at the moment of creation, synchronized to a network time server. Time zones are recorded with UTC offset.
- **Original:** The first record of an observation is the original record. PX-BMS never modifies original records — changes create new records with the original preserved in the audit trail.

- **Accurate:** Records must reflect what actually happened. Proc-X instrument calibration program and instrument error monitoring ensure that recorded process parameters accurately represent actual process conditions.
- **Complete, Consistent, Enduring, Available:** All data from every batch is archived indefinitely. Full data sets are available for regulatory inspection on demand. No selective data archiving.

28.3 21 CFR Part 11 Requirements for Electronic Records

21 CFR Part 11 applies to electronic records and electronic signatures used to satisfy a requirement or regulation administered by the FDA. For pharmaceutical processing equipment, the relevant records include: batch production records (21 CFR 211.188), equipment use logs (211.182), calibration records (211.68), and cleaning records (211.67). Part 11 requirements:

- **Validation:** Systems must be validated to ensure accuracy, reliability, consistent intended performance, and the ability to discern invalid or altered records.
- **Audit trail:** Systems must use computer-generated, time-stamped audit trails to independently record the date and time of operator entries and actions that create, modify, or delete electronic records.
- **Access controls:** Systems must limit access to authorized individuals via use of passwords, biometric devices, or other authority checks.
- **Electronic signatures:** Must link the signer's name to the specific record, include the date and time of signing, and include the meaning of the signature.

28.4 Common Data Integrity Failures in Processing Equipment

Failure Mode	How It Occurs	Proc-X Prevention
Shared generic login	Operators share 'admin' or 'operator' password	Named user accounts only; shared accounts architecturally impossible
Disabled audit trail	System administrator disables audit trail to save storage space	Audit trail cannot be disabled by any user level; storage auto-expands
Manual data transcription	Process data printed and hand-entered into LIMS or EBR	PX-BMS exports directly to LIMS/EBR via OPC-UA; no manual transcription
Clock manipulation	System clock changed to backdate records	Hardware clock synchronized to network time server; clock change requires dual authorization and creates audit event
Selective printing	Only favorable batch data printed; rest deleted	All data retained in historian; print function prints complete batch record only
Test/practice entries in production system	Operators use production system for training	Separate training environment provisioned on all Proc-X installations

Chapter 29: Continuous Manufacturing and ICH Q13

IN THIS CHAPTER

- ▶ What makes continuous manufacturing different from batch — and why regulators now support it
- ▶ ICH Q13 requirements: control strategy, RTD, diversion, and batch definition
- ▶ The complete continuous solid oral dose manufacturing line
- ▶ Real-time release testing (RTRT) — the quality system that enables continuous manufacturing
- ▶ PAT instrument network architecture for continuous pharmaceutical manufacturing

29.1 Continuous Manufacturing — the Paradigm Shift

Continuous pharmaceutical manufacturing (CM) produces drug product in a continuously flowing process rather than discrete batches. Raw materials are fed continuously into the process and finished product is collected continuously at the other end, with the entire process running at steady state. The business case for CM is compelling: smaller equipment footprint (because throughput is governed by time × flow rate, not vessel size), faster production (hours instead of days per batch cycle), reduced in-process inventory, lower energy consumption, and — critically — better process control through real-time monitoring of every unit of product.

The regulatory acceptance of CM has transformed from skepticism (pre-2012) to active encouragement. The FDA's 2019 guidance on adopting pharmaceutical continuous manufacturing, followed by ICH Q13 in 2022, provides a clear regulatory pathway for CM submissions in new drug applications (NDAs) and abbreviated new drug applications (ANDAs).

29.2 The Complete Continuous Solid Dose Line

Process Step	Equipment	Key PAT Instruments	Control Strategy
Continuous feeding	LIW feeders (API + excipients)	Gravimetric mass flow, Raman (API ID)	Closed-loop feed rate control; API ID confirmation
Continuous blending	Continuous tube blender or plow blender	NIR (blend uniformity, API content)	NIR feedback to feeder speed; endpoint by MBSD
Continuous granulation	TSE wet granulator	FBRM (granule size), NIR (moisture)	Liquid-to-solid ratio control; spray rate cascade
Continuous drying	Fluid bed dryer (continuous)	NIR (moisture), product temp	Spray rate / air flow cascade; NIR endpoint
Continuous milling	Comil or ACM	Inline PSD (laser backscatter)	Classifier speed feedback to PSD target
Continuous blending (final)	Continuous V-blender	NIR (content uniformity)	API feeder cascade; diversion on out-of-spec

Continuous compression	Rotary tablet press	Weight + hardness (in-press monitoring)	Force feedback; automatic diversion on weight OOS
Continuous coating	Perforated drum coater (continuous)	NIR (film weight gain)	Spray rate + pan speed for film weight target

29.3 Real-Time Release Testing (RTRT)

RTRT replaces end-product testing with real-time monitoring of surrogate process attributes that have been shown to predict end-product quality. The regulatory basis is ICH Q8 (pharmaceutical development) and the FDA's 2004 PAT guidance. For a continuous solid dose line with RTRT:

- Content uniformity: NIR blend uniformity monitoring replaces the 30-unit USP <905> end-product test. ICH Q13 requires demonstration that the NIR model predicts CU acceptance value with appropriate statistical confidence.
- Dissolution: NIR physical characterization of granule surface area and particle size can predict dissolution in some formulations. Where not possible, rapid in-line dissolution testing (fiber optic UV cell in a compact dissolution vessel) is used.
- Identity: Raman or NIR spectroscopy in the blend stream provides continuous identity confirmation, replacing periodic offline HPLC identification testing.

PROC-X EQUIPMENT SPOTLIGHT Proc-X continuous manufacturing lines are supplied with the PX-RTRT data package: a multivariate calibration model library for NIR/Raman-based content uniformity and moisture monitoring, pre-built for the most common pharmaceutical excipient systems (MCC/HPMC/magnesium stearate and lactose/starch/PVP). Custom model development for novel formulations is available through Proc-X's pharmaceutical science team.

Chapter 30: The Digital Pharmaceutical Facility — Data, AI, and the Future

IN THIS CHAPTER

- ▶ Industry 4.0 in pharmaceutical manufacturing: the promise and the regulatory constraints
- ▶ The pharmaceutical digital twin: simulating process performance before it happens
- ▶ Machine learning for process optimization and failure prediction
- ▶ The fully connected facility: OPC-UA, ISA-88/95, and the pharma data backbone

30.1 Industry 4.0 and Pharma's Unique Regulatory Context

Industry 4.0 — the integration of cyber-physical systems, Internet of Things connectivity, advanced analytics, and artificial intelligence in manufacturing — is transforming pharmaceutical manufacturing. But the pharmaceutical context adds a regulatory dimension that distinguishes it from every other industry: any algorithmic system used to make a GMP decision must be validated under GAMP 5 or an equivalent framework. An AI model that adjusts a critical process parameter based on PAT data is a regulated computer system — it requires qualification, validation, periodic review, and change control.

The FDA's 2021 discussion paper on AI and machine learning in drug manufacturing acknowledges this challenge while encouraging innovation. The paper establishes that AI/ML systems used in GMP contexts are subject to Part 11 and GAMP 5 requirements, and that the validation framework for these systems must address the additional complexity of models that may drift or change performance over time.

30.2 The Pharmaceutical Digital Twin

A digital twin of a pharmaceutical manufacturing process is a continuously updated mathematical model that represents the current state of the physical process in real time. For a Proc-X-equipped granulation line, the digital twin would incorporate: first-principles models of high-shear granule growth kinetics, fluid bed drying mass transfer, and blender mixing kinetics; calibrated against historical batch data from the PX-BMS historian; and updated in real time with incoming PAT data from the production line. The digital twin enables:

- Predictive scale-up: Simulate the effect of batch size change before committing production equipment.
- Process troubleshooting: Identify root cause of batch deviation by simulating alternative process histories.
- Optimization: Find the parameter combination that minimizes cycle time while maintaining all CQA specifications within the design space.

30.3 Machine Learning for Pharmaceutical Processing

- Batch failure prediction: LSTM recurrent neural networks trained on PX-BMS historian data predict tablet hardness, content uniformity, and dissolution from mid-process granulation parameters — enabling early batch diversion decisions before downstream processing costs are incurred.
- Predictive maintenance: Vibration and temperature monitoring of gearboxes, impeller drives, and fluid bed fan motors combined with ML anomaly detection identifies bearing and seal failures 2–6 weeks before they cause production shutdowns.
- Formulation optimization: Gaussian process regression models in the design of experiments framework accelerate formulation screening by predicting the optimal binder concentration and granulation endpoint for new APIs from a limited set of characterization data.

REGULATORY INSIGHT *The FDA's CDER (Center for Drug Evaluation and Research) has received over 200 submissions containing real-time release testing components since 2006, and the number is accelerating with the adoption of continuous manufacturing. The message from these submissions is clear: regulators are fully prepared to accept PAT and RTRT when the analytical method is validated, the process model is sound, and the data integrity systems are robust. The barrier to RTRT adoption is not regulatory — it is the investment in method development and system validation that most manufacturers have been reluctant to make. Proc-X's RTRT support packages are designed to eliminate this barrier.*

APPENDIX A: Glossary of Key Terms

ALCOA+: Data integrity framework: Attributable, Legible, Contemporaneous, Original, Accurate, plus Complete, Consistent, Enduring, and Available.

Amorphous Solid Dispersion (ASD): A solid pharmaceutical form in which the API is molecularly dispersed in a polymer matrix in an amorphous (non-crystalline) state, producing enhanced dissolution rate.

API: Active Pharmaceutical Ingredient. The pharmacologically active molecule in a drug product.

BCS: Biopharmaceutics Classification System. Classifies APIs by aqueous solubility (high/low) and intestinal permeability (high/low). BCS Class II (high permeability, low solubility) requires particle size reduction for bioavailability.

cGMP: Current Good Manufacturing Practice. The minimum standard required for manufacturing, processing, packing, and holding of drugs. Governed in the U.S. by 21 CFR Parts 210 and 211.

CIP: Clean-in-Place. A method of cleaning pharmaceutical equipment in place without disassembly, using automated spray systems and cleaning solution circulation.

CPP: Critical Process Parameter. A process parameter whose variability has an impact on a critical quality attribute and therefore should be monitored or controlled to ensure the process produces the desired quality.

CQA: Critical Quality Attribute. A physical, chemical, biological, or microbiological property or characteristic that should be within an appropriate limit, range, or distribution to ensure the desired product quality.

CPV: Continued Process Verification. Stage 3 of the FDA Process Validation lifecycle — ongoing statistical monitoring of process performance in commercial manufacturing.

Design Space: The multidimensional combination and interaction of input variables (material attributes) and process parameters that have been demonstrated to provide assurance of quality. Defined in ICH Q8.

DQ: Design Qualification. Documents that equipment/system has been designed to meet defined requirements.

FBP: Fluid Bed Processor. Equipment in which heated air is forced upward through a powder bed to suspend it in a fluidized state for granulation, drying, or coating.

GAMP 5: Good Automated Manufacturing Practice, 5th edition. Industry guidance for validation of automated systems in pharmaceutical manufacturing, published by ISPE.

HME: Hot-Melt Extrusion. A process in which API and polymer are melted and extruded to form an amorphous solid dispersion for bioavailability enhancement.

HPAPI: High-Potency Active Pharmaceutical Ingredient. An API with an occupational exposure limit $\leq 10 \mu\text{g}/\text{m}^3$ (OEB 4) or $\leq 1 \mu\text{g}/\text{m}^3$ (OEB 5).

ICH Q13: International Council for Harmonisation guideline on Continuous Manufacturing of Drug Substances and Drug Products (2022).

IQ: Installation Qualification. Documented verification that equipment has been installed correctly and consistently with supplier recommendations and industry standards.

LIW Feeder: Loss-in-Weight Feeder. A gravimetric continuous feeder that measures material feed rate by monitoring the rate of weight loss from the feeder hopper.

MMAD: Mass Median Aerodynamic Diameter. The aerodynamic particle diameter at which 50% of the inhaled mass is larger and 50% is smaller. Target MMAD for DPI lung deposition: 1–5 μm .

OEB: Occupational Exposure Band. Classification system for API potency and required containment level: OEB 1 (lowest potency) through OEB 5 (highest potency).

OQ: Operational Qualification. Documented verification that equipment operates per its specifications across its operating range.

PAT: Process Analytical Technology. A system for designing, analyzing, and controlling manufacturing through timely measurements of critical quality and performance attributes of raw and in-process materials.

PQ: Performance Qualification. Documented verification that the process and equipment consistently produce a product meeting all critical quality attributes.

PX-BMS: Proc-X Batch Management System. The SCADA, historian, and electronic batch record system integrated into all Proc-X pharmaceutical processing lines.

QbD: Quality by Design. A systematic approach to pharmaceutical development beginning with predefined objectives and emphasizing product and process understanding based on sound science and quality risk management.

TRT: Real-Time Release Testing. Evaluation and certification of an in-process or final product based on process data, which provides greater assurance of product quality than end-product testing.

SIP: Sterilize-in-Place. Sterilization of a vessel or system in situ using saturated steam at 121°C for ≥ 30 minutes without dismantling.

TSE: Twin-Screw Extruder. A co-rotating or counter-rotating twin-screw processing machine used in pharmaceutical HME, continuous wet granulation, and polymer compounding.

21 CFR Part 11: Title 21 CFR Part 11. FDA regulation establishing criteria under which the agency considers electronic records and electronic signatures to be trustworthy, reliable, and equivalent to paper records.

21 CFR Part 211: Current Good Manufacturing Practice for Finished Pharmaceuticals. The primary U.S. regulatory standard governing pharmaceutical manufacturing equipment, facilities, and processes.

APPENDIX B: Equipment Selection Guide

Process Task	Material / Application	Recommended Equipment	Key Process Parameter
Wet granulation (batch)	All solid dose formulations	PX-HSG High-Shear Granulator	Impeller tip speed, L/S ratio, power endpoint
Dry granulation	Moisture-sensitive APIs, low-dose	PX-RC Roller Compactor + Comil	Roll pressure, ribbon density, screen aperture
Fluid bed granulation	Free-flowing powders, top-spray	PX-FBP (top-spray insert)	Inlet temp, spray rate, product temp endpoint
Fluid bed drying	Wet granules from HSG or TSG	PX-FBP or PX-ONE (one-pot)	Inlet temp, product temp, LOD / NIR endpoint
Wurster coating	Pellets, granules, tablets (ER/enteric)	PX-FBP (Wurster insert)	Spray rate, coat weight gain, partition gap
Pan coating (tablets)	Immediate release, taste masking	PX-COAT side-vented pan	Pan speed, spray rate, inlet air, exhaust RH
API micronization (D90 < 50 µm)	BCS Class II APIs, DPI formulations	PX-API Jet Mill + N2 system	Classifier speed, grinding pressure
API nanosizing (< 1 µm)	BCS Class II oral suspension, injectable	Wet bead mill (0.1 mm ZrO2 media)	Milling time / passes, bead fill, temperature
Cryogenic milling	Thermolabile APIs, amorphous ASDs	Proc-X cryo jet mill or cryo pin mill	LN2 feed rate, mill zone temperature
HME (bioavailability enhancement)	BCS Class II ASD production	PX-HME Twin-Screw Extruder	Barrel temp profile, screw speed, throughput
Continuous wet granulation	High-throughput continuous solid dose	PX-CWG TSE + FBP dryer	Liquid addition rate, screw speed, FBRM endpoint
Bin blending (potent APIs)	OEB 3–5 blending, multi-component blends	PX-IBC Bin Blender + contained charging	Rotations, tip speed, NIR blend endpoint
Sterile mixing	Parenterals, lyophilized injectables	PX-SV Sterile Vessel (magnetic drive)	Mixing speed, time, temperature, conductivity
Softgel fill preparation	Oil-based nutraceuticals, lipophilic APIs	High-shear disperser + vacuum de-aeration	Shear rate, de-aeration time, dissolved O2
Contained powder transfer	OEB 4/5 HPAPIs	PX-ISO containment isolator	OEL verification (air sampling), glove integrity
Continuous API feeding	Continuous manufacturing lines	PX-LIW Loss-in-Weight Feeder	Feed rate accuracy, refill transition algorithm

APPENDIX C: Process Parameter Reference Tables

C.1 High-Shear Wet Granulation Reference Parameters

Binder System	Concentration	Addition Rate	Wet Massing Speed	Power Endpoint
PVP K30 (aqueous)	5–10% w/w in water	10–50 mL/min/kg powder	High (600 rpm impeller)	1.2–1.5× dry blend power
HPMC E5 (aqueous)	3–8% w/w in water	8–40 mL/min/kg	Medium (400 rpm)	1.3–1.6× dry blend power
Hydroxypropyl starch	5–15% w/w in water	10–60 mL/min/kg	Medium (400 rpm)	1.2–1.5× dry blend power
Ethanol (dry binder solvent)	95% ethanol, no dissolved binder	5–20 mL/min/kg	High (500–600 rpm)	Visible wet mass formation
Aqueous dispersion (Aquacoat ECG)	15% solids dispersion	5–15 mL/min/kg	Low–medium (300 rpm)	By granule PSD offline

C.2 Jet Milling Reference Parameters for Common API Classes

API Class	Target D50 (µm)	Classifier (rpm)	N2 Pressure (psig)	Feed Rate (kg/hr)
BCS Class II oral	5–20	5,000–10,000	90–110	5–50
DPI inhalation API	1.5–3.5	11,000–14,000	110–130	0.5–5
Low-dose (< 1% w/w)	5–15	6,000–10,000	90–110	1–10
High-potency (OEB 4)	3–15	6,000–12,000	90–120	0.5–20
Thermolabile (cryogenic)	5–25	5,000–9,000	80–100	2–20
Nutraceutical botanical	10–50	3,000–7,000	80–100	10–200

C.3 Fluid Bed Drying Reference Parameters

Granule Type	Inlet Temp (°C)	Product Temp (°C)	LOD Target (%)	Typical Drying Time
Standard aqueous granulation	60–80	35–50	1.0–3.0%	30–90 min
Moisture-sensitive API	40–60	25–40	0.5–2.0%	60–150 min
High-dose (> 70% API)	55–70	40–55	1.0–2.5%	45–120 min
Ethanol granulation	50–70 (solvent-rated FBP)	35–45	0.5–1.5%	20–60 min
Sustained release pellets	50–65	35–48	1.0–2.5%	60–120 min

APPENDIX D: Regulatory Standards Index

Standard	Title	Issuing Body
21 CFR Part 210	Current GMP in Manufacturing, Processing, Packing, or Holding of Drugs — General	FDA
21 CFR Part 211	Current GMP for Finished Pharmaceuticals	FDA
21 CFR Part 111	Current GMP in Manufacturing, Packaging, Labeling, or Holding Operations for Dietary Supplements	FDA
21 CFR Part 11	Electronic Records; Electronic Signatures	FDA
21 CFR Part 820	Quality System Regulation for Medical Devices	FDA
EU GMP Part I	Basic Requirements for Medicinal Products	EMA
EU GMP Annex 11	Computerised Systems	EMA
EU GMP Annex 1 (2022)	Manufacture of Sterile Medicinal Products	EMA
EU GMP Annex 15	Qualification and Validation	EMA
ICH Q7	Good Manufacturing Practice for Active Pharmaceutical Ingredients	ICH
ICH Q8(R2)	Pharmaceutical Development (QbD Framework)	ICH
ICH Q9(R1)	Quality Risk Management	ICH
ICH Q10	Pharmaceutical Quality System	ICH
ICH Q11	Development and Manufacture of Drug Substances	ICH
ICH Q13	Continuous Manufacturing of Drug Substances and Drug Products	ICH
USP <1>	Injections and Implanted Drug Products — General	USP
USP <429>	Light Diffraction Measurement of Particle Size	USP
USP <601>	Aerosols, Nasal Sprays, Metered-Dose Inhalers, and Dry Powder Inhalers	USP
USP <711>	Dissolution	USP
USP <905>	Uniformity of Dosage Units	USP
Ph. Eur. 2.9.31	Particle Size Analysis by Laser Light Diffraction	Ph. Eur.
Ph. Eur. 2.9.18	Preparations for Inhalation: Aerodynamic Assessment of Fine Particles	Ph. Eur.
GAMP 5	Good Automated Manufacturing Practice, 5th Edition	ISPE
PIC/S PI 006	Recommendations on Validation Master Plan, IQ, OQ, NQ, PQ	PIC/S
ISO 14644-1	Cleanrooms and Associated Controlled Environments — Classification of Air Cleanliness	ISO
ASTM A270	Standard Specification for Seamless and Welded Austenitic	ASTM

	Stainless Steel Tubing for General Service	
ASTM A967	Standard Specification for Chemical Passivation Treatments for Stainless Steel Parts	ASTM

APPENDIX E: Proc-X Pharmaceutical Equipment Quick-Reference Index

Model Series	Equipment Type	Size Range	Primary Application	Key cGMP Feature
PX-HSG	High-Shear Granulator	2–600 L	Wet granulation, all solid oral dose	Power/torque logging; gravimetric binder control
PX-RC	Roller Compactor	Lab–production	Dry granulation, moisture-sensitive APIs	Gap control ± 0.01 mm; inline NIR ribbon density
PX-FBP	Fluid Bed Processor	5–400 kg	Granulation, drying, Wurster coating	Modular inserts; inline NIR LOD; cascade control
PX-ONE	One-Pot Processor	10–150 L	Combined HSG + fluid bed drying	Single vessel eliminates transfer contamination risk
PX-COAT	Side-Vented Coating Pan	10–600 kg	IR, ER, enteric, taste-mask coating	IR exhaust temp control; inline NIR film weight
PX-API	Pharmaceutical Jet Mill	100 g–500 kg/hr	API micronization, DPI particle sizing	N2 system; OEB 4/5 containment; inline PSD
PX-HME	Hot-Melt Extruder	11–40 mm screw	ASD production for BCS Class II APIs	Zone temp uniformity $\pm 1^\circ\text{C}$; 21 CFR Part 11 logging
PX-CWG	Continuous Wet Granulation Line	5–150 kg/hr	Continuous solid dose manufacturing	Integrated FBRM + NIR + FBP dryer + PX-BMS
PX-IBC	Bin (IBC) Blender	50–1,500 L	Potent compound blending OEB 3–5	Contained charging; RFID IBC tracking; NIR blend endpoint
PX-SV	Sterile Mixing Vessel	50–10,000 L	Parenteral compounding, lyo preparation	Magnetic drive (no shaft seal); full SIP; $R_a \leq 0.5 \mu\text{m}$
PX-LIW	Loss-in-Weight Feeder	0.1–5,000 g/hr	Continuous manufacturing API/excipient feeding	Fast-refill bulk density compensation; Raman API ID
PX-BMS	Batch Management System	1–200 nodes	All Proc-X pharmaceutical systems	GAMP 5 Cat. 4; 21 CFR Part 11; OPC-UA; RTRT-ready



THE FORMULATION ENGINEER

cGMP Processing Equipment for
**Pharmaceutical & Nutraceutical
Manufacturing**

Mixers · Mills · Extruders · Bulk Material Handling

Engineering compliance into every machine.

www.procxinc.com | john.paul@procxinc.com

© 2026 Proc-X Manufacturing Group. All rights reserved.
No portion of this document may be reproduced without written permission.